

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX101018\
 Data File : VX005215.D
 Acq On : 10 Oct 2018 19:07
 Operator : JC/MD
 Sample : J5317-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-109D1-20181005

Quant Time: Oct 11 07:17:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	173897	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	275601	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	272537	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	151264	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	133096	53.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.16%	
35) Dibromofluoromethane	5.51	113	112385	48.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.24%	
50) Toluene-d8	8.72	98	401122	51.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.32%	
62) 4-Bromofluorobenzene	11.14	95	142462	50.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.86%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.39	101	1443	0.76	ug/l	98
16) Acetone	2.45	43	6064	4.18	ug/l	93
44) Trichloroethene	7.21	130	65097	26.54	ug/l	95
64) Tetrachloroethene	9.34	164	1164	0.43	ug/l	96
95) Naphthalene	13.83	128	1311	0.82	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX101018\
Data File : VX005215.D
Acq On : 10 Oct 2018 19:07
Operator : JC/MD
Sample : J5317-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-109D1-20181005

Quant Time: Oct 11 07:17:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

