

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX101018\
 Data File : VX005218.D
 Acq On : 10 Oct 2018 20:28
 Operator : JC/MD
 Sample : J5317-10 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-108D2-20181004

Quant Time: Oct 11 07:39:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	167120	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	263314	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	262076	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	144595	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	130043	54.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.94%	
35) Dibromofluoromethane	5.51	113	109193	48.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.86%	
50) Toluene-d8	8.72	98	387265	52.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.40%	
62) 4-Bromofluorobenzene	11.14	95	135160	50.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.14%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
44) Trichloroethene	7.21	130	101039	43.12	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX101018\
Data File : VX005218.D
Acq On : 10 Oct 2018 20:28
Operator : JC/MD
Sample : J5317-10 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-108D2-20181004

Quant Time: Oct 11 07:39:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

