

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX101018\
 Data File : VX005223.D
 Acq On : 10 Oct 2018 22:42
 Operator : JC/MD
 Sample : J5317-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-108D1-20181004

Quant Time: Oct 11 07:48:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	167790	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	267958	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	270814	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	149193	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	133302	55.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.22%	
35) Dibromofluoromethane	5.50	113	112060	49.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.68%	
50) Toluene-d8	8.72	98	391029	51.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.58%	
62) 4-Bromofluorobenzene	11.14	95	139989	51.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.94%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	3791	2.71	ug/l	# 86
44) Trichloroethene	7.21	130	59893	25.12	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX101018\
Data File : VX005223.D
Acq On : 10 Oct 2018 22:42
Operator : JC/MD
Sample : J5317-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-108D1-20181004

Quant Time: Oct 11 07:48:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

