

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX101018\
 Data File : VX005224.D
 Acq On : 10 Oct 2018 23:09
 Operator : JC/MD
 Sample : J5317-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-109D2-20181005

Quant Time: Oct 11 07:50:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	163173	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	260531	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	259419	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	144616	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	128807	55.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.52%	
35) Dibromofluoromethane	5.51	113	109637	49.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.30%	
50) Toluene-d8	8.72	98	381638	51.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.98%	
62) 4-Bromofluorobenzene	11.14	95	135027	51.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.12%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	1634	0.92	ug/l	99
16) Acetone	2.45	43	7608	5.59	ug/l	99
44) Trichloroethene	7.21	130	74671	32.21	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX101018\
Data File : VX005224.D
Acq On : 10 Oct 2018 23:09
Operator : JC/MD
Sample : J5317-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-109D2-20181005

Quant Time: Oct 11 07:50:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

