

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX101018\
 Data File : VX005235.D
 Acq On : 11 Oct 2018 05:51
 Operator : JC/MD
 Sample : J5284-08DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-120D2-20181002DL

Quant Time: Oct 11 08:28:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	169595	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	259967	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	264144	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	148496	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	131201	54.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.32%	
35) Dibromofluoromethane	5.50	113	111494	50.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.20%	
50) Toluene-d8	8.72	98	383320	52.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.66%	
62) 4-Bromofluorobenzene	11.14	95	136551	51.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.50%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.38	101	2372	1.28	ug/l	97
44) Trichloroethene	7.21	130	121366	52.46	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX101018\
Data File : VX005235.D
Acq On : 11 Oct 2018 05:51
Operator : JC/MD
Sample : J5284-08DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-120D2-20181002DL

Quant Time: Oct 11 08:28:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

