

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX101018\
 Data File : VX005238.D
 Acq On : 11 Oct 2018 07:12
 Operator : JC/MD
 Sample : J5284-11DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-103D2-20181003DL

Quant Time: Oct 11 08:33:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	167937	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	255017	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	254780	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	138460	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	127091	52.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.96%	
35) Dibromofluoromethane	5.51	113	107433	49.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.42%	
50) Toluene-d8	8.72	98	372831	51.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.78%	
62) 4-Bromofluorobenzene	11.14	95	130930	50.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.16%	
Target Compounds						
44) Trichloroethene	7.21	130	112390	49.52	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX101018\
Data File : VX005238.D
Acq On : 11 Oct 2018 07:12
Operator : JC/MD
Sample : J5284-11DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-103D2-20181003DL

Quant Time: Oct 11 08:33:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

