

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX101018\
 Data File : VX005244.D
 Acq On : 11 Oct 2018 09:53
 Operator : JC/MD
 Sample : J5364-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY038PS038-181008

Quant Time: Oct 11 12:06:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	180863	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	270715	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	269546	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	147846	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	135452	52.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.86%	
35) Dibromofluoromethane	5.50	113	115798	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
50) Toluene-d8	8.72	98	395422	51.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.68%	
62) 4-Bromofluorobenzene	11.14	95	138214	50.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.60%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	15881	10.53	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX101018\
Data File : VX005244.D
Acq On : 11 Oct 2018 09:53
Operator : JC/MD
Sample : J5364-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY038PS038-181008

Quant Time: Oct 11 12:06:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

