

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX101018\
 Data File : VX005246.D
 Acq On : 11 Oct 2018 10:47
 Operator : JC/MD
 Sample : J5390-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA18-CHY07

Quant Time: Oct 12 09:56:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	167124	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	263441	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	268212	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	154802	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	130220	54.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.10%	
35) Dibromofluoromethane	5.51	113	112237	50.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.54%	
50) Toluene-d8	8.72	98	385193	51.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.78%	
62) 4-Bromofluorobenzene	11.14	95	140178	52.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.84%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	41935	30.09	ug/l	93
17) Carbon Disulfide	2.57	76	2577	0.47	ug/l #	94
18) Methyl Acetate	2.78	43	8867	2.51	ug/l	99
25) 2-Butanone	4.71	43	5502	2.68	ug/l	99
27) cis-1,2-Dichloroethene	4.60	96	4502	2.03	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX101018\
Data File : VX005246.D
Acq On : 11 Oct 2018 10:47
Operator : JC/MD
Sample : J5390-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA18-CHY07

Quant Time: Oct 12 09:56:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

