

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101019\
 Data File : VX012983.D
 Acq On : 10 Oct 2019 10:12
 Operator : JC/SP
 Sample : VX1010WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1010WBL01

Quant Time: Oct 11 05:04:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	179699	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	300953	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	256647	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	100368	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	114961	51.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.36%	
35) Dibromofluoromethane	5.48	113	90575	52.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.88%	
50) Toluene-d8	8.71	98	350185	51.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.96%	
62) 4-Bromofluorobenzene	11.14	95	121597	46.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.90%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101019\
Data File : VX012983.D
Acq On : 10 Oct 2019 10:12
Operator : JC/SP
Sample : VX1010WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1010WBL01

Quant Time: Oct 11 05:04:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

