

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101019\
 Data File : VX012987.D
 Acq On : 10 Oct 2019 11:50
 Operator : JC/SP
 Sample : K5213-09DL 100X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 W-1DL

Quant Time: Oct 11 05:57:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	183227	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	303341	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	259987	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	101939	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	117197	51.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.36%	
35) Dibromofluoromethane	5.49	113	89909	51.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.28%	
50) Toluene-d8	8.71	98	358220	52.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.50%	
62) 4-Bromofluorobenzene	11.14	95	123012	46.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.24%	
Target Compounds						
4) Vinyl Chloride	1.40	62	112517	53.632	ug/l	98
27) cis-1,2-Dichloroethene	4.59	96	197708	86.426	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101019\
Data File : VX012987.D
Acq On : 10 Oct 2019 11:50
Operator : JC/SP
Sample : K5213-09DL 100X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
W-1DL

Quant Time: Oct 11 05:57:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

