

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101019\
 Data File : VX012988.D
 Acq On : 10 Oct 2019 12:14
 Operator : JC/SP
 Sample : K5261-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-TB-2019-9

Quant Time: Oct 11 06:25:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	175952	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	294133	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	251120	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	98165	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	114572	52.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.20%	
35) Dibromofluoromethane	5.49	113	87708	51.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.90%	
50) Toluene-d8	8.71	98	344246	52.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.56%	
62) 4-Bromofluorobenzene	11.14	95	114406	44.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.44%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101019\
Data File : VX012988.D
Acq On : 10 Oct 2019 12:14
Operator : JC/SP
Sample : K5261-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S49-TB-2019-9

Quant Time: Oct 11 06:25:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

