

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101019\
 Data File : VX012991.D
 Acq On : 10 Oct 2019 13:24
 Operator : JC/SP
 Sample : K5291-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-JMW1881

Quant Time: Oct 11 06:44:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	182142	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	301840	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	259640	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	100426	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	118921	52.24	ug/l	0.00
Spiked Amount						
			Recovery	=	104.48%	
35) Dibromofluoromethane	5.49	113	90239	52.09	ug/l	0.00
Spiked Amount						
			Recovery	=	104.18%	
50) Toluene-d8	8.71	98	356455	52.76	ug/l	0.00
Spiked Amount						
			Recovery	=	105.52%	
62) 4-Bromofluorobenzene	11.14	95	121222	46.17	ug/l	0.00
Spiked Amount						
			Recovery	=	92.34%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
3) Chloromethane	1.32	50	1430	0.708	ug/l	99
4) Vinyl Chloride	1.40	62	10854	5.204	ug/l	98
16) Acetone	2.44	43	7699	6.416	ug/l	95
21) trans-1,2-Dichloroethene	3.16	96	1335	0.677	ug/l #	81
27) cis-1,2-Dichloroethene	4.59	96	39143	17.213	ug/l	91
37) Ethyl Acetate	4.83	43	15767	5.313	ug/l	99
40) Benzene	6.14	78	4086	0.521	ug/l #	78
44) Trichloroethene	7.20	130	5741	2.641	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101019\
Data File : VX012991.D
Acq On : 10 Oct 2019 13:24
Operator : JC/SP
Sample : K5291-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-JMW1881

Quant Time: Oct 11 06:44:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

