

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101019\
 Data File : VX012995.D
 Acq On : 10 Oct 2019 14:57
 Operator : JC/SP
 Sample : K5290-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-MMW0125R

Quant Time: Oct 11 07:18:17 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	177579	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	294053	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	253007	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	97953	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	114163	51.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.88%	
35) Dibromofluoromethane	5.49	113	89254	52.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.76%	
50) Toluene-d8	8.71	98	350327	53.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.44%	
62) 4-Bromofluorobenzene	11.14	95	115527	45.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.34%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.58	96	1957	0.883	ug/l	94
44) Trichloroethene	7.20	130	2108	0.996	ug/l	77
64) Tetrachloroethene	9.33	164	80144	45.670	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101019\
Data File : VX012995.D
Acq On : 10 Oct 2019 14:57
Operator : JC/SP
Sample : K5290-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-MMW0125R

Quant Time: Oct 11 07:18:17 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

