

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101019\
 Data File : VX012998.D
 Acq On : 10 Oct 2019 16:07
 Operator : JC/SP
 Sample : K5290-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FA19-JMW3601

Quant Time: Oct 11 07:34:34 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 09 01:51:23 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	168296	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	284067	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	240986	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	87420	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	110705	52.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.26%	
35) Dibromofluoromethane	5.49	113	84895	52.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.14%	
50) Toluene-d8	8.71	98	331014	52.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.10%	
62) 4-Bromofluorobenzene	11.14	95	107754	43.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.22%	
Target Compounds						
16) Acetone	2.44	43	1943	1.752	ug/l	96
44) Trichloroethene	7.20	130	2870	1.403	ug/l	92
64) Tetrachloroethene	9.33	164	6588	3.941	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101019\
Data File : VX012998.D
Acq On : 10 Oct 2019 16:07
Operator : JC/SP
Sample : K5290-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FA19-JMW3601

Quant Time: Oct 11 07:34:34 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 09 01:51:23 2019
Response via : Initial Calibration

