

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101022\
 Data File : VX032000.D
 Acq On : 10 Oct 2022 10:21
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Quant Time: Oct 11 05:05:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	64637	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	98166	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	105585	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75654	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67437	50.519	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 101.040%	
35) Dibromofluoromethane	5.385	113	57164	51.444	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.880%	
50) Toluene-d8	8.647	98	200673	50.188	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.380%	
62) 4-Bromofluorobenzene	11.079	95	75870	53.673	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 107.340%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	51702	53.747	ug/l	99
3) Chloromethane	1.288	50	46357	46.020	ug/l	97
4) Vinyl Chloride	1.374	62	54826	46.573	ug/l	98
5) Bromomethane	1.611	94	37937	49.685	ug/l	97
6) Chloroethane	1.685	64	51039	50.949	ug/l	93
7) Trichlorofluoromethane	1.886	101	102949	48.652	ug/l	99
8) Diethyl Ether	2.130	74	35498	49.215	ug/l	94
9) 1,1,2-Trichlorotrifluo...	2.325	101	52939	49.496	ug/l	92
10) Methyl Iodide	2.453	142	52799	48.597	ug/l	100
11) Tert butyl alcohol	2.989	59	37841	227.844	ug/l	97
12) 1,1-Dichloroethene	2.319	96	47338	48.081	ug/l	90
13) Acrolein	2.233	56	27925	294.397	ug/l	100
14) Allyl chloride	2.660	41	71964	51.370	ug/l	97
15) Acrylonitrile	3.062	53	128886	243.435	ug/l	99
16) Acetone	2.386	43	105784	233.169	ug/l	95
17) Carbon Disulfide	2.508	76	105882	50.080	ug/l	98
18) Methyl Acetate	2.703	43	76064	50.814	ug/l	94
19) Methyl tert-butyl Ether	3.111	73	171366	51.868	ug/l	98
20) Methylene Chloride	2.788	84	54974	46.112	ug/l	92
21) trans-1,2-Dichloroethene	3.087	96	52589	48.233	ug/l	95
22) Diisopropyl ether	3.757	45	168099	53.376	ug/l	99
23) Vinyl Acetate	3.715	43	667086	283.376	ug/l	98
24) 1,1-Dichloroethane	3.605	63	94278	49.940	ug/l	97
25) 2-Butanone	4.556	43	175530	247.503	ug/l	99
26) 2,2-Dichloropropane	4.471	77	73694	54.347	ug/l	96
27) cis-1,2-Dichloroethene	4.483	96	62693	48.055	ug/l	91
28) Bromochloromethane	4.891	49	40675	50.337	ug/l	90
29) Tetrahydrofuran	5.001	42	109904	240.178	ug/l	95
30) Chloroform	5.093	83	107212	50.985	ug/l	97
31) Cyclohexane	5.471	56	78835	49.436	ug/l	95
32) 1,1,1-Trichloroethane	5.379	97	91296	51.091	ug/l	99
36) 1,1-Dichloropropene	5.690	75	73561	49.923	ug/l	99
37) Ethyl Acetate	4.708	43	71114	51.263	ug/l	99
38) Carbon Tetrachloride	5.672	117	81395	54.025	ug/l	98
39) Methylcyclohexane	7.379	83	92643	52.711	ug/l	94
40) Benzene	6.031	78	216655	50.213	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101022\
 Data File : VX032000.D
 Acq On : 10 Oct 2022 10:21
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050

Quant Time: Oct 11 05:05:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	39728	55.073	ug/l	96
42) 1,2-Dichloroethane	6.086	62	84252	53.586	ug/l	97
43) Isopropyl Acetate	6.336	43	117814	55.976	ug/l	98
44) Trichloroethene	7.123	130	61587	48.680	ug/l	93
45) 1,2-Dichloropropane	7.428	63	56360	53.152	ug/l	95
46) Dibromomethane	7.580	93	42826	52.115	ug/l	93
47) Bromodichloromethane	7.818	83	84202	59.314	ug/l	98
48) Methyl methacrylate	7.690	41	56908	56.583	ug/l	93
49) 1,4-Dioxane	7.696	88	22067	967.681	ug/l	98
51) 4-Methyl-2-Pentanone	8.574	43	369498	264.915	ug/l	98
52) Toluene	8.714	92	142065	50.358	ug/l	100
53) t-1,3-Dichloropropene	8.976	75	82707	54.022	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	91764	59.689	ug/l	96
55) 1,1,2-Trichloroethane	9.153	97	59850	50.265	ug/l	99
56) Ethyl methacrylate	9.116	69	84844	55.590	ug/l	97
57) 1,3-Dichloropropane	9.305	76	99264	51.879	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.238	63	205526	260.870	ug/l	95
59) 2-Hexanone	9.433	43	269888	262.428	ug/l	98
60) Dibromochloromethane	9.519	129	65268	59.897	ug/l	99
61) 1,2-Dibromoethane	9.610	107	61848	51.652	ug/l	98
64) Tetrachloroethene	9.275	164	60653	50.882	ug/l	99
65) Chlorobenzene	10.079	112	154257	48.217	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	60410	54.571	ug/l	99
67) Ethyl Benzene	10.195	91	276896	51.161	ug/l	99
68) m/p-Xylenes	10.299	106	216710	100.760	ug/l	100
69) o-Xylene	10.640	106	106133	50.770	ug/l	100
70) Styrene	10.653	104	178609	53.446	ug/l	99
71) Bromoform	10.799	173	44062	60.801	ug/l #	99
73) Isopropylbenzene	10.963	105	284061	49.783	ug/l	98
74) N-amyl acetate	10.842	43	103475	59.384	ug/l	98
75) 1,1,2,2-Tetrachloroethane	11.213	83	86798	47.877	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	93726	60.448	ug/l	85
77) Bromobenzene	11.201	156	68464	48.710	ug/l	97
78) n-propylbenzene	11.305	91	332863	51.288	ug/l	99
79) 2-Chlorotoluene	11.366	91	193646	49.581	ug/l	99
80) 1,3,5-Trimethylbenzene	11.451	105	242942	51.336	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	23094	52.648	ug/l	90
82) 4-Chlorotoluene	11.457	91	227259	50.807	ug/l	99
83) tert-Butylbenzene	11.713	119	234818	50.134	ug/l	97
84) 1,2,4-Trimethylbenzene	11.750	105	242544	50.451	ug/l	99
85) sec-Butylbenzene	11.890	105	305756	52.728	ug/l	99
86) p-Isopropyltoluene	12.006	119	263317	53.933	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	130986	48.945	ug/l	97
88) 1,4-Dichlorobenzene	12.043	146	133838	48.587	ug/l	99
89) n-Butylbenzene	12.335	91	231548	56.633	ug/l	99
90) Hexachloroethane	12.536	117	43304	55.627	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	131890	49.568	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	18116	54.590	ug/l	89
93) 1,2,4-Trichlorobenzene	13.591	180	86087	53.307	ug/l	100
94) Hexachlorobutadiene	13.725	225	37932	58.074	ug/l	98
95) Naphthalene	13.774	128	279594	50.525	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	85220	52.024	ug/l	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101022\
Data File : VX032000.D
Acq On : 10 Oct 2022 10:21
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Quant Time: Oct 11 05:05:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:168 Resp: 64637
Ion Ratio Lower Upper
168 100
99 54.0 48.8 73.2






#2
Dichlorodifluoromethane
Concen: 53.747 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion: 85 Resp: 51702
Ion Ratio Lower Upper
85 100
87 32.3 15.8 47.4



Abundance Scan 33 (1.288 min): VX031915.D\data.ms (-29)



#3

Chloromethane

Concen: 46.020 ug/l

RT: 1.288 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX032000.D

Acq: 10 Oct 2022 10:21

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050

Tgt Ion: 50 Resp: 46357




Ion Ratio Lower Upper

50 100

52 32.0 26.9 40.3



Abundance Scan 47 (1.373 min): VX031915.D\data.ms (-41)



#4

Vinyl Chloride

Concen: 46.573 ug/l

RT: 1.374 min Scan# 47

Delta R.T. 0.001 min

Lab File: VX032000.D


Acq: 10 Oct 2022 10:21


Tgt Ion: 62 Resp: 54826

Ion Ratio Lower Upper

62 100

64 30.6 25.3 37.9









#5
Bromomethane
Concen: 49.685 ug/l
RT: 1.611 min Scan# 81
Delta R.T. 0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 94 Resp: 37937
Ion Ratio Lower Upper
94 100
96 96.1 74.4 111.6



#6
Chloroethane
Concen: 50.949 ug/l
RT: 1.685 min Scan# 98
Delta R.T. 0.001 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion: 64 Resp: 51039
Ion Ratio Lower Upper
64 100
66 35.0 25.0 37.4









#7
Trichlorofluoromethane
Concen: 48.652 ug/l
RT: 1.886 min Scan# 11
Delta R.T. 0.001 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:101 Resp: 102949
Ion Ratio Lower Upper
101 100
103 64.1 51.8 77.6



#8
Diethyl Ether
Concen: 49.215 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.005 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion: 74 Resp: 35498
Ion Ratio Lower Upper
74 100
45 91.3 48.5 145.6









#9
1,1,2-Trichlorotrifluoroethane
Concen: 49.496 ug/l
RT: 2.325 min Scan# 204
Delta R.T. -0.006 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:101 Resp: 52939
Ion Ratio Lower Upper
101 100
85 46.4 38.6 57.8
151 76.8 54.0 81.0



#10
Methyl Iodide
Concen: 48.597 ug/l
RT: 2.453 min Scan# 224
Delta R.T. 0.001 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion:142 Resp: 52799
Ion Ratio Lower Upper
142 100
127 46.6 37.4 56.0
141 14.6 11.6 17.4









#11
Tert butyl alcohol
Concen: 227.844 ug/l
RT: 2.989 min Scan# 315
Delta R.T. -0.018 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 59 Resp: 37841
Ion Ratio Lower Upper
59 100
57 9.1 8.2 12.2



#12
1,1-Dichloroethene
Concen: 48.081 ug/l
RT: 2.319 min Scan# 202
Delta R.T. 0.001 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion: 96 Resp: 47338
Ion Ratio Lower Upper
96 100
61 151.1 135.3 202.9
98 62.0 49.4 74.2









#13
Acrolein
Concen: 294.397 ug/l
RT: 2.233 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 56 Resp: 27925
Ion Ratio Lower Upper
56 100
55 70.2 56.0 84.0



#14
Allyl chloride
Concen: 51.370 ug/l
RT: 2.660 min Scan# 258
Delta R.T. -0.006 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion: 41 Resp: 71964
Ion Ratio Lower Upper
41 100
39 73.6 58.2 87.4
76 38.2 26.9 40.3









#15
Acrylonitrile
Concen: 243.435 ug/l
RT: 3.062 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 53 Resp: 128886
Ion Ratio Lower Upper
53 100
52 82.8 66.9 100.3
51 37.4 30.0 45.0



#16
Acetone
Concen: 233.169 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.006 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion: 43 Resp: 105784
Ion Ratio Lower Upper
43 100
58 32.3 23.8 35.8









#17
Carbon Disulfide
Concen: 50.080 ug/l
RT: 2.508 min Scan# 21
Delta R.T. -0.005 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 76 Resp: 105882
Ion Ratio Lower Upper
76 100
78 9.6 7.2 10.8



#18
Methyl Acetate
Concen: 50.814 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.006 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion: 43 Resp: 76064
Ion Ratio Lower Upper
43 100
74 26.1 18.6 27.8





#19

Methyl tert-butyl Ether

Concen: 51.868 ug/l

RT: 3.111 min Scan# 31

Delta R.T. -0.006 min

Lab File: VX032000.D

Acq: 10 Oct 2022 10:21

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 73 Resp: 171366

Ion Ratio Lower Upper

73 100

57 20.7 17.2 25.8



#20

Methylene Chloride

Concen: 46.112 ug/l

RT: 2.788 min Scan# 279

Delta R.T. 0.000 min

Lab File: VX032000.D

Acq: 10 Oct 2022 10:21

Tgt Ion: 84 Resp: 54974




Ion Ratio Lower Upper


84 100

49 119.6 105.9 158.9

51 36.8 32.2 48.2

86 66.5 50.8 76.2









#21
trans-1,2-Dichloroethene
Concen: 48.233 ug/l
RT: 3.087 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 96 Resp: 52589
Ion Ratio Lower Upper
96 100
61 127.7 106.7 160.1
98 65.1 48.7 73.1



#22
Diisopropyl ether
Concen: 53.376 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.006 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion: 45 Resp: 168099
Ion Ratio Lower Upper
45 100
43 74.4 59.0 88.6
87 29.2 22.6 33.8
59 11.4 8.8 13.2









#23
 Vinyl Acetate
 Concen: 283.376 ug/l
 RT: 3.715 min Scan# 413
 Delta R.T. -0.006 min
 Lab File: VX032000.D
 Acq: 10 Oct 2022 10:21




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion: 43 Resp: 667086
 Ion Ratio Lower Upper
 43 100
 86 11.5 8.7 13.1



#24
 1,1-Dichloroethane
 Concen: 49.940 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.006 min
 Lab File: VX032000.D
 Acq: 10 Oct 2022 10:21

Tgt Ion: 63 Resp: 94278
 Ion Ratio Lower Upper
 63 100
 98 8.7 3.5 10.4
 100 4.7 2.4 7.2





#25

2-Butanone

Concen: 247.503 ug/l

RT: 4.556 min Scan# 5

Delta R.T. -0.006 min

Lab File: VX032000.D

Acq: 10 Oct 2022 10:21

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 43 Resp: 175530

Ion Ratio Lower Upper

43 100

72 27.0 21.1 31.7



#26

2,2-Dichloropropane

Concen: 54.347 ug/l

RT: 4.471 min Scan# 555

Delta R.T. -0.005 min

Lab File: VX032000.D




Acq: 10 Oct 2022 10:21


Tgt Ion: 77 Resp: 73694

Ion Ratio Lower Upper

77 100

97 23.4 10.9 32.6









#27
cis-1,2-Dichloroethene
Concen: 48.055 ug/l
RT: 4.483 min Scan# 51
Delta R.T. -0.006 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 96 Resp: 62693
Ion Ratio Lower Upper
96 100
61 137.2 0.0 305.2
98 65.0 0.0 128.4



#28
Bromochloromethane
Concen: 50.337 ug/l
RT: 4.891 min Scan# 624
Delta R.T. -0.006 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion: 49 Resp: 40675
Ion Ratio Lower Upper
49 100
129 2.1 0.0 3.6
130 82.6 59.0 88.4





#29

Tetrahydrofuran

Concen: 240.178 ug/l

RT: 5.001 min Scan# 644

Delta R.T. -0.012 min

Lab File: VX032000.D

Acq: 10 Oct 2022 10:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 42 Resp: 109904

Ion Ratio Lower Upper

42 100

72 48.4 36.4 54.6

71 45.4 33.5 50.3



#30

Chloroform

Concen: 50.985 ug/l

RT: 5.093 min Scan# 657

Delta R.T. 0.001 min

Lab File: VX032000.D




Acq: 10 Oct 2022 10:21


Tgt Ion: 83 Resp: 107212

Ion Ratio Lower Upper

83 100

85 68.3 52.8 79.2









#31
Cyclohexane
Concen: 49.436 ug/l
RT: 5.471 min Scan# 719
Delta R.T. 0.001 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 56 Resp: 78835
Ion Ratio Lower Upper
56 100
69 26.9 25.1 37.7
84 91.2 70.2 105.4



#32
1,1,1-Trichloroethane
Concen: 51.091 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.006 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion: 97 Resp: 91296
Ion Ratio Lower Upper
97 100
99 65.1 51.3 76.9
61 45.4 35.9 53.9









#33
1,2-Dichloroethane-d4
Concen: 50.519 ug/l
RT: 5.952 min Scan# 799
Delta R.T. -0.006 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 65 Resp: 67437
Ion Ratio Lower Upper
65 100
67 51.8 0.0 102.4



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 930
Delta R.T. -0.006 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion: 114 Resp: 98166
Ion Ratio Lower Upper
114 100
63 19.8 0.0 42.6
88 15.2 0.0 33.0









#35
Dibromofluoromethane
Concen: 51.444 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:113 Resp: 57164
Ion Ratio Lower Upper
113 100
111 102.3 82.4 123.6
192 19.1 14.1 21.1



#36
1,1-Dichloropropene
Concen: 49.923 ug/l
RT: 5.690 min Scan# 755
Delta R.T. 0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion: 75 Resp: 73561
Ion Ratio Lower Upper
75 100
110 36.4 17.5 52.5
77 31.1 24.9 37.3





#37

Ethyl Acetate

Concen: 51.263 ug/l

RT: 4.708 min Scan# 594

Delta R.T. -0.006 min

Lab File: VX032000.D

Acq: 10 Oct 2022 10:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 43 Resp: 71114

Ion Ratio Lower Upper

43 100

61 14.5 11.1 16.7

70 11.4 8.6 13.0



#38

Carbon Tetrachloride

Concen: 54.025 ug/l

RT: 5.672 min Scan# 752

Delta R.T. -0.005 min

Lab File: VX032000.D

Acq: 10 Oct 2022 10:21




Tgt Ion: 117 Resp: 81395


Ion Ratio Lower Upper

117 100

119 96.3 74.6 112.0

121 31.3 25.3 37.9





#39

Methylcyclohexane

Concen: 52.711 ug/l

RT: 7.379 min Scan# 1032

Delta R.T. -0.006 min

Lab File: VX032000.D

Acq: 10 Oct 2022 10:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 83 Resp: 92643

Ion Ratio Lower Upper

83 100

55 69.5 61.0 91.4

98 49.9 37.7 56.5



#40

Benzene

Concen: 50.213 ug/l

RT: 6.031 min Scan# 811

Delta R.T. -0.006 min

Lab File: VX032000.D




Acq: 10 Oct 2022 10:21


Tgt Ion: 78 Resp: 216655

Ion Ratio Lower Upper

78 100

77 23.9 18.7 28.1









#41
Methacrylonitrile
Concen: 55.073 ug/l
RT: 4.916 min Scan# 61
Delta R.T. -0.006 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 41 Resp: 39728
Ion Ratio Lower Upper
41 100
39 56.8 45.0 67.6
67 77.7 58.2 87.4
52 33.1 25.0 37.6



#42
1,2-Dichloroethane
Concen: 53.586 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.006 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion: 62 Resp: 84252
Ion Ratio Lower Upper
62 100
98 9.9 0.0 17.8





#43

Isopropyl Acetate

Concen: 55.976 ug/l

RT: 6.336 min Scan# 80

Delta R.T. -0.006 min

Lab File: VX032000.D

Acq: 10 Oct 2022 10:21

Instrument :





MSVOA_X

ClientSampleId :

VSTDCCC050

Tgt Ion: 43 Resp: 117814

Ion	Ratio	Lower	Upper
43	100		
61	19.3	15.3	22.9
87	14.5	10.6	15.8



#44

Trichloroethene

Concen: 48.680 ug/l

RT: 7.123 min Scan# 990




Delta R.T. -0.005 min


Lab File: VX032000.D

Acq: 10 Oct 2022 10:21

Tgt Ion: 130 Resp: 61587

Ion	Ratio	Lower	Upper
130	100		
95	95.9	0.0	206.8





#45

1,2-Dichloropropane

Concen: 53.152 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. -0.005 min

Lab File: VX032000.D

Acq: 10 Oct 2022 10:21

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 63 Resp: 56360

Ion Ratio Lower Upper

63 100

65 30.3 26.6 39.8



#46

Dibromomethane

Concen: 52.115 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.000 min

Lab File: VX032000.D

Acq: 10 Oct 2022 10:21




Tgt Ion: 93 Resp: 42826


Ion Ratio Lower Upper

93 100

95 82.4 65.5 98.3

174 96.9 67.6 101.4









#47
Bromodichloromethane
Concen: 59.314 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.005 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 84202
Ion Ratio Lower Upper
83 100
85 61.9 51.0 76.6
127 8.3 7.0 10.6



#48
Methyl methacrylate
Concen: 56.583 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. -0.005 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion: 41 Resp: 56908
Ion Ratio Lower Upper
41 100
69 89.2 64.6 96.8
39 66.7 51.4 77.0





#49

1,4-Dioxane

Concen: 967.681 ug/l

RT: 7.696 min Scan# 1084

Delta R.T. 0.001 min

Lab File: VX032000.D

Acq: 10 Oct 2022 10:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion: 88 Resp: 22067

Ion Ratio Lower Upper

88 100

43 33.4 27.1 40.7

58 65.8 53.9 80.9



#50

Toluene-d8

Concen: 50.188 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.006 min

Lab File: VX032000.D




Acq: 10 Oct 2022 10:21


Tgt Ion: 98 Resp: 200673

Ion Ratio Lower Upper

98 100

100 66.6 52.5 78.7





#51

4-Methyl-2-Pentanone

Concen: 264.915 ug/l

RT: 8.574 min Scan# 11

Delta R.T. 0.001 min

Lab File: VX032000.D

Acq: 10 Oct 2022 10:21

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 43 Resp: 369498

Ion Ratio Lower Upper

43 100

58 39.3 30.6 46.0



#52

Toluene

Concen: 50.358 ug/l

RT: 8.714 min Scan# 1251

Delta R.T. -0.006 min

Lab File: VX032000.D




Acq: 10 Oct 2022 10:21


Tgt Ion: 92 Resp: 142065

Ion Ratio Lower Upper

92 100

91 171.2 137.1 205.7





#53

t-1,3-Dichloropropene

Concen: 54.022 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.006 min

Lab File: VX032000.D

Acq: 10 Oct 2022 10:21

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 75 Resp: 82707

Ion Ratio Lower Upper

75 100

77 31.0 24.8 37.2



#54

cis-1,3-Dichloropropene

Concen: 59.689 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX032000.D

Acq: 10 Oct 2022 10:21




Tgt Ion: 75 Resp: 91764


Ion Ratio Lower Upper

75 100

77 33.0 24.4 36.6

39 51.5 43.7 65.5









#55
1,1,2-Trichloroethane
Concen: 50.265 ug/l
RT: 9.153 min Scan# 11
Delta R.T. 0.001 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 97 Resp: 59850




Ion	Ratio	Lower	Upper
97	100		
83	85.8	69.0	103.6
85	53.7	44.1	66.1
99	64.5	51.3	76.9




#56
Ethyl methacrylate
Concen: 55.590 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion: 69 Resp: 84844

Ion	Ratio	Lower	Upper
69	100		
41	66.2	54.2	81.4
39	44.8	34.2	51.2









#57
1,3-Dichloropropane
Concen: 51.879 ug/l
RT: 9.305 min Scan# 1173
Delta R.T. -0.006 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 76 Resp: 99264
Ion Ratio Lower Upper
76 100
78 32.1 25.5 38.3



#58
2-Chloroethyl Vinyl ether
Concen: 260.870 ug/l
RT: 8.238 min Scan# 1173
Delta R.T. -0.006 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion: 63 Resp: 205526
Ion Ratio Lower Upper
63 100
106 30.7 22.6 34.0





#59

2-Hexanone

Concen: 262.428 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX032000.D

Acq: 10 Oct 2022 10:21

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 43 Resp: 269888

Ion Ratio Lower Upper

43 100

58 54.8 26.7 80.0



#60

Dibromochloromethane

Concen: 59.897 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. -0.005 min

Lab File: VX032000.D




Acq: 10 Oct 2022 10:21


Tgt Ion: 129 Resp: 65268

Ion Ratio Lower Upper

129 100

127 76.6 38.7 116.1









#61
1,2-Dibromoethane
Concen: 51.652 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:107 Resp: 61848
Ion Ratio Lower Upper
107 100
109 95.2 74.2 111.4



#62
4-Bromofluorobenzene
Concen: 53.673 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion: 95 Resp: 75870
Ion Ratio Lower Upper
95 100
174 80.3 0.0 153.2
176 77.7 0.0 144.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:117 Resp: 105585
Ion Ratio Lower Upper
117 100
82 53.5 46.8 70.2
119 30.3 26.4 39.6



#64
Tetrachloroethene
Concen: 50.882 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. 0.001 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion:164 Resp: 60653
Ion Ratio Lower Upper
164 100
166 126.3 99.0 148.4
129 97.7 78.0 117.0
131 92.8 74.2 111.2









#65
Chlorobenzene
Concen: 48.217 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:112 Resp: 154257
Ion Ratio Lower Upper
112 100
114 31.3 25.7 38.5



#66
1,1,1,2-Tetrachloroethane
Concen: 54.571 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.005 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21


Tgt Ion:131 Resp: 60410
Ion Ratio Lower Upper
131 100
133 92.2 46.6 139.7
119 63.4 31.9 95.9








#67
Ethyl Benzene
Concen: 51.161 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050







Tgt Ion: 91 Resp: 276896
Ion Ratio Lower Upper
91 100
106 31.1 25.2 37.8



#68
m/p-Xylenes
Concen: 100.760 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.006 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion: 106 Resp: 216710
Ion Ratio Lower Upper
106 100
91 201.7 161.0 241.4









#69
o-Xylene
Concen: 50.770 ug/l
RT: 10.640 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:106 Resp: 106133
Ion Ratio Lower Upper
106 100
91 215.1 107.2 321.6



#70
Styrene
Concen: 53.446 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. -0.005 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion:104 Resp: 178609
Ion Ratio Lower Upper
104 100
78 50.6 41.4 62.2
103 54.6 43.8 65.8









#71
Bromoform
Concen: 60.801 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:173 Resp: 44062
Ion Ratio Lower Upper
173 100
175 48.8 24.0 72.0
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion:152 Resp: 75654
Ion Ratio Lower Upper
152 100
115 88.4 43.4 130.1
150 174.7 0.0 346.8









#73
Isopropylbenzene
Concen: 49.783 ug/l
RT: 10.963 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:105 Resp: 284061
Ion Ratio Lower Upper
105 100
120 26.5 13.7 41.0



#74
N-ethyl acetate
Concen: 59.384 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. -0.005 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion: 43 Resp: 103475
Ion Ratio Lower Upper
43 100
70 42.7 33.4 50.2
55 24.3 18.2 27.2
61 23.8 19.3 28.9









#75
1,1,2,2-Tetrachloroethane
Concen: 47.877 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 83 Resp: 86798
Ion Ratio Lower Upper
83 100
131 10.5 5.1 15.4
85 64.4 32.6 97.8



#76
1,2,3-Trichloropropane
Concen: 60.448 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. 0.001 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion: 75 Resp: 93726
Ion Ratio Lower Upper
75 100
77 31.1 20.3 60.9









#77
Bromobenzene
Concen: 48.710 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:156 Resp: 68464
Ion Ratio Lower Upper
156 100
77 154.1 79.5 238.6
158 95.5 48.4 145.3



#78
n-propylbenzene
Concen: 51.288 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion: 91 Resp: 332863
Ion Ratio Lower Upper
91 100
120 23.2 11.8 35.4





#79

2-Chlorotoluene

Concen: 49.581 ug/l

RT: 11.366 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX032000.D

Acq: 10 Oct 2022 10:21

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050

Tgt Ion: 91 Resp: 193646

Ion Ratio Lower Upper

91 100

126 34.4 17.4 52.3



#80

1,3,5-Trimethylbenzene

Concen: 51.336 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX032000.D




Acq: 10 Oct 2022 10:21


Tgt Ion: 105 Resp: 242942

Ion Ratio Lower Upper

105 100

120 49.4 24.5 73.5









#81
trans-1,4-Dichloro-2-butene
Concen: 52.648 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion: 75 Resp: 23094
Ion Ratio Lower Upper
75 100
53 111.7 77.6 116.4
89 47.0 37.4 56.0



#82
4-Chlorotoluene
Concen: 50.807 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion: 91 Resp: 227259
Ion Ratio Lower Upper
91 100
126 29.8 15.1 45.3









#83
tert-Butylbenzene
Concen: 50.134 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. 0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:119 Resp: 234818
Ion Ratio Lower Upper
119 100
91 61.2 29.2 87.6
134 23.6 11.6 34.8



#84
1,2,4-Trimethylbenzene
Concen: 50.451 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.006 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion:105 Resp: 242544
Ion Ratio Lower Upper
105 100
120 44.8 22.6 67.8









#85
 sec-Butylbenzene
 Concen: 52.728 ug/l
 RT: 11.890 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX032000.D
 Acq: 10 Oct 2022 10:21




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050


Tgt Ion:105 Resp: 305756
 Ion Ratio Lower Upper
 105 100
 134 20.2 10.4 31.1



#86
 p-Isopropyltoluene
 Concen: 53.933 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.006 min
 Lab File: VX032000.D
 Acq: 10 Oct 2022 10:21

Tgt Ion:119 Resp: 263317
 Ion Ratio Lower Upper
 119 100
 134 26.3 13.4 40.2
 91 25.2 12.4 37.2





#87

1,3-Dichlorobenzene

Concen: 48.945 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX032000.D

Acq: 10 Oct 2022 10:21

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050





Tgt Ion:146 Resp: 130986

Ion Ratio Lower Upper

146 100

111 42.1 19.8 59.4

148 64.4 31.5 94.5



#88

1,4-Dichlorobenzene

Concen: 48.587 ug/l

RT: 12.043 min Scan# 1797

Delta R.T. 0.001 min

Lab File: VX032000.D

Acq: 10 Oct 2022 10:21




Tgt Ion:146 Resp: 133838


Ion Ratio Lower Upper

146 100

111 40.8 20.5 61.5

148 63.0 32.0 96.2









#89
n-Butylbenzene
Concen: 56.633 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion: 91 Resp: 231548




Ion	Ratio	Lower	Upper
91	100		
92	53.2	27.0	81.0
134	26.2	13.3	39.9




#90
Hexachloroethane
Concen: 55.627 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.006 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion: 117 Resp: 43304

Ion	Ratio	Lower	Upper
117	100		
201	68.7	34.8	104.4









#91
1,2-Dichlorobenzene
Concen: 49.568 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:146 Resp: 131890
Ion Ratio Lower Upper
146 100
111 42.3 20.6 61.8
148 64.1 31.6 95.0



#92
1,2-Dibromo-3-Chloropropane
Concen: 54.590 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.001 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion: 75 Resp: 18116
Ion Ratio Lower Upper
75 100
155 88.6 40.3 120.8
157 115.7 51.5 154.5









#93
1,2,4-Trichlorobenzene
Concen: 53.307 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050


Tgt Ion:180 Resp: 86087
Ion Ratio Lower Upper
180 100
182 93.4 46.9 140.6
145 33.0 16.7 50.1



#94
Hexachlorobutadiene
Concen: 58.074 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion:225 Resp: 37932
Ion Ratio Lower Upper
225 100
223 64.0 30.8 92.4
227 63.2 32.3 96.9









#95
Naphthalene
Concen: 50.525 ug/l
RT: 13.774 min Scan# 2082
Delta R.T. -0.006 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050

Tgt Ion:128 Resp: 279594
Ion Ratio Lower Upper
128 100
127 13.0 10.4 15.6
129 10.7 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 52.024 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX032000.D
Acq: 10 Oct 2022 10:21

Tgt Ion:180 Resp: 85220
Ion Ratio Lower Upper
180 100
182 95.8 47.9 143.8
145 33.3 17.6 52.8

