

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101022\
 Data File : VX032015.D
 Acq On : 10 Oct 2022 16:17
 Operator : JC/MD
 Sample : N5054-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 510A-1

Quant Time: Oct 11 05:12:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	67638	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	104049	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	108745	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	66541	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69088	49.459	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.920%
35) Dibromofluoromethane	5.385	113	56182	47.702	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.400%
50) Toluene-d8	8.653	98	209897	49.527	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.060%
62) 4-Bromofluorobenzene	11.079	95	69339	46.280	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.560%
Target Compounds						
16) Acetone	2.392	43	738	1.555	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101022\
Data File : VX032015.D
Acq On : 10 Oct 2022 16:17
Operator : JC/MD
Sample : N5054-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
510A-1

Quant Time: Oct 11 05:12:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

