

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101022\
 Data File : VX032019.D
 Acq On : 10 Oct 2022 17:50
 Operator : JC/MD
 Sample : N4937-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-104D

Quant Time: Oct 11 06:00:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	47054	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	79874	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	84287	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	51153	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	56358	57.996	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 116.000%			
35) Dibromofluoromethane	5.385	113	45910	50.778	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.560%			
50) Toluene-d8	8.646	98	161192	49.546	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.100%			
62) 4-Bromofluorobenzene	11.079	95	54064	47.006	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 94.020%			
Target Compounds					Qvalue	
16) Acetone	2.392	43	855	2.589	ug/l	92
19) Methyl tert-butyl Ether	3.117	73	2017	0.839	ug/l	94
30) Chloroform	5.086	83	4033	2.635	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101022\
Data File : VX032019.D
Acq On : 10 Oct 2022 17:50
Operator : JC/MD
Sample : N4937-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-104D

Quant Time: Oct 11 06:00:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

