

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101022\
 Data File : VX032020.D
 Acq On : 10 Oct 2022 18:13
 Operator : JC/MD
 Sample : N4937-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-114D

Quant Time: Oct 11 06:00:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 06 14:28:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	50924	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	83337	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	88969	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	53527	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	59955	57.008	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.020%
35) Dibromofluoromethane	5.385	113	48417	51.326	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.660%
50) Toluene-d8	8.653	98	168013	49.497	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.000%
62) 4-Bromofluorobenzene	11.079	95	56311	46.925	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.860%
Target Compounds						
						Qvalue
19) Methyl tert-butyl Ether	3.117	73	2180	0.838	ug/l #	87
20) Methylene Chloride	2.788	84	2430	2.587	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101022\
Data File : VX032020.D
Acq On : 10 Oct 2022 18:13
Operator : JC/MD
Sample : N4937-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-114D

Quant Time: Oct 11 06:00:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 06 14:28:46 2022
Response via : Initial Calibration

