

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101023\
 Data File : VX038213.D
 Acq On : 10 Oct 2023 19:47
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 11 01:43:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	131473	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	220502	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	222221	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	129908	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	99578	47.089	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.180%
35) Dibromofluoromethane	5.385	113	78943	49.685	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.380%
50) Toluene-d8	8.647	98	305269	50.280	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.560%
62) 4-Bromofluorobenzene	11.079	95	120434	52.323	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.640%

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	67079	52.858	ug/l	98
3) Chloromethane	1.301	50	64300	49.712	ug/l	100
4) Vinyl Chloride	1.374	62	74507	51.954	ug/l	99
5) Bromomethane	1.599	94	106369	56.587	ug/l	99
6) Chloroethane	1.679	64	61763	57.996	ug/l	97
7) Trichlorofluoromethane	1.886	101	156399	55.213	ug/l	97
8) Diethyl Ether	2.130	74	57175	54.121	ug/l	80
9) 1,1,2-Trichlorotrifluo...	2.325	101	64681	52.306	ug/l	97
10) Methyl Iodide	2.453	142	90407	50.812	ug/l	95
11) Tert butyl alcohol	2.959	59	107775	229.992	ug/l #	92
12) 1,1-Dichloroethene	2.319	96	62396	51.872	ug/l	90
13) Acrolein	2.233	56	99566	331.201	ug/l	99
14) Allyl chloride	2.660	41	105326	48.954	ug/l	94
15) Acrylonitrile	3.062	53	243343	252.867	ug/l	99
16) Acetone	2.380	43	216845	245.702	ug/l	92
17) Carbon Disulfide	2.514	76	156601	45.171	ug/l	98
18) Methyl Acetate	2.703	43	196685	51.969	ug/l	93
19) Methyl tert-butyl Ether	3.111	73	243093	50.743	ug/l	100
20) Methylene Chloride	2.788	84	77727	46.822	ug/l	91
21) trans-1,2-Dichloroethene	3.093	96	74571	50.070	ug/l	89
22) Diisopropyl ether	3.757	45	224928	49.881	ug/l #	96
23) Vinyl Acetate	3.721	43	836590	252.987	ug/l	96
24) 1,1-Dichloroethane	3.611	63	136544	49.896	ug/l	100
25) 2-Butanone	4.556	43	352557	247.587	ug/l #	90
26) 2,2-Dichloropropane	4.477	77	100631	47.099	ug/l	96
27) cis-1,2-Dichloroethene	4.489	96	89058	50.209	ug/l	91
28) Bromochloromethane	4.897	49	65092	52.690	ug/l	88
29) Tetrahydrofuran	5.001	42	228102	246.039	ug/l	89
30) Chloroform	5.093	83	155398	51.929	ug/l	100
31) Cyclohexane	5.477	56	110417	48.194	ug/l	94
32) 1,1,1-Trichloroethane	5.385	97	132653	51.753	ug/l	99
36) 1,1-Dichloropropene	5.690	75	108140	49.356	ug/l	98
37) Ethyl Acetate	4.715	43	132213	50.810	ug/l	97
38) Carbon Tetrachloride	5.678	117	113179	52.510	ug/l	95
39) Methylcyclohexane	7.379	83	126590	49.327	ug/l	93
40) Benzene	6.038	78	316301	51.614	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101023\
 Data File : VX038213.D
 Acq On : 10 Oct 2023 19:47
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 11 01:43:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.916	41	68636	53.023	ug/l	91
42) 1,2-Dichloroethane	6.086	62	124565	50.876	ug/l	99
43) Isopropyl Acetate	6.342	43	199511	51.360	ug/l	97
44) Trichloroethene	7.129	130	83599	52.158	ug/l	97
45) 1,2-Dichloropropane	7.428	63	82129	51.781	ug/l	98
46) Dibromomethane	7.580	93	63170	51.352	ug/l	97
47) Bromodichloromethane	7.824	83	118451	54.128	ug/l	98
48) Methyl methacrylate	7.690	41	96927	52.268	ug/l #	85
49) 1,4-Dioxane	7.659	88	45857	1003.506	ug/l #	88
51) 4-Methyl-2-Pentanone	8.574	43	716830	267.535	ug/l	97
52) Toluene	8.720	92	213118	53.388	ug/l	93
53) t-1,3-Dichloropropene	8.976	75	128653	54.769	ug/l	100
54) cis-1,3-Dichloropropene	8.366	75	131959	53.469	ug/l	92
55) 1,1,2-Trichloroethane	9.153	97	91620	53.659	ug/l	95
56) Ethyl methacrylate	9.116	69	142154	54.301	ug/l	92
57) 1,3-Dichloropropane	9.311	76	148087	53.624	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	321569	250.244	ug/l	99
59) 2-Hexanone	9.427	43	575990	267.401	ug/l	95
60) Dibromochloromethane	9.519	129	95536	57.170	ug/l	100
61) 1,2-Dibromoethane	9.610	107	100617	54.492	ug/l	100
64) Tetrachloroethene	9.275	164	73518	51.104	ug/l	98
65) Chlorobenzene	10.079	112	235130	50.565	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	87518	55.658	ug/l	99
67) Ethyl Benzene	10.195	91	427806	51.541	ug/l	98
68) m/p-Xylenes	10.299	106	333394	103.879	ug/l	98
69) o-Xylene	10.640	106	161277	51.487	ug/l	100
70) Styrene	10.653	104	271890	53.299	ug/l	97
71) Bromoform	10.799	173	70650	48.523	ug/l #	98
73) Isopropylbenzene	10.963	105	430690	49.436	ug/l	99
74) N-amyl acetate	10.842	43	185403	49.835	ug/l	97
75) 1,1,2,2-Tetrachloroethane	11.213	83	160587	50.234	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	172760	58.497	ug/l	91
77) Bromobenzene	11.195	156	105612	49.654	ug/l	91
78) n-propylbenzene	11.305	91	521787	48.982	ug/l	100
79) 2-Chlorotoluene	11.366	91	302754	47.923	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	375987	50.024	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.018	75	44081	44.101	ug/l	94
82) 4-Chlorotoluene	11.451	91	362584	48.502	ug/l	97
83) tert-Butylbenzene	11.713	119	375735	51.160	ug/l	96
84) 1,2,4-Trimethylbenzene	11.750	105	376188	50.198	ug/l	98
85) sec-Butylbenzene	11.890	105	480551	50.512	ug/l	100
86) p-Isopropyltoluene	12.006	119	400103	50.973	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	207600	48.594	ug/l	98
88) 1,4-Dichlorobenzene	12.043	146	211612	48.518	ug/l	98
89) n-Butylbenzene	12.335	91	385306	49.290	ug/l	97
90) Hexachloroethane	12.536	117	67727	52.041	ug/l	93
91) 1,2-Dichlorobenzene	12.335	146	204872	49.890	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.939	75	38706	50.077	ug/l	86
93) 1,2,4-Trichlorobenzene	13.585	180	132722	48.188	ug/l	97
94) Hexachlorobutadiene	13.725	225	53686	48.429	ug/l	99
95) Naphthalene	13.774	128	462655	48.235	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	129480	48.551	ug/l	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101023\
Data File : VX038213.D
Acq On : 10 Oct 2023 19:47
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Oct 11 01:43:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101023\
 Data File : VX038213.D
 Acq On : 10 Oct 2023 19:47
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :


Quant Time: Oct 11 01:43:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:168 Resp: 131473
Ion Ratio Lower Upper
168 100
99 58.3 56.6 85.0



#2
Dichlorodifluoromethane
Concen: 52.858 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47

Tgt Ion: 85 Resp: 67079
Ion Ratio Lower Upper
85 100
87 32.5 16.7 50.1





#3

Chloromethane

Concen: 49.712 ug/l

RT: 1.301 min Scan# 35

Delta R.T. -0.000 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 50 Resp: 64300

Ion Ratio Lower Upper

50 100

52 32.2 25.5 38.3



#4

Vinyl Chloride

Concen: 51.954 ug/l

RT: 1.374 min Scan# 47

Delta R.T. -0.000 min

Lab File: VX038213.D




Acq: 10 Oct 2023 19:47


Tgt Ion: 62 Resp: 74507

Ion Ratio Lower Upper

62 100

64 32.2 25.4 38.0









#5
 Bromomethane
 Concen: 56.587 ug/l
 RT: 1.599 min Scan# 84
 Delta R.T. 0.000 min
 Lab File: VX038213.D
 Acq: 10 Oct 2023 19:47




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 94 Resp: 106369
 Ion Ratio Lower Upper
 94 100
 96 94.5 74.6 111.8



#6
 Chloroethane
 Concen: 57.996 ug/l
 RT: 1.679 min Scan# 97
 Delta R.T. -0.000 min
 Lab File: VX038213.D
 Acq: 10 Oct 2023 19:47

Tgt Ion: 64 Resp: 61763
 Ion Ratio Lower Upper
 64 100
 66 32.8 25.1 37.7









#7
 Trichlorofluoromethane
 Concen: 55.213 ug/l
 RT: 1.886 min Scan# 131
 Delta R.T. -0.000 min
 Lab File: VX038213.D
 Acq: 10 Oct 2023 19:47




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:101 Resp: 156399
 Ion Ratio Lower Upper
 101 100
 103 66.9 51.4 77.2



#8
 Diethyl Ether
 Concen: 54.121 ug/l
 RT: 2.130 min Scan# 171
 Delta R.T. -0.000 min
 Lab File: VX038213.D
 Acq: 10 Oct 2023 19:47

Tgt Ion: 74 Resp: 57175
 Ion Ratio Lower Upper
 74 100
 45 85.4 53.3 159.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 52.306 ug/l
RT: 2.325 min Scan# 203
Delta R.T. -0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 64681
Ion Ratio Lower Upper
101 100
85 47.3 37.0 55.6
151 76.5 58.6 87.8



#10
Methyl Iodide
Concen: 50.812 ug/l
RT: 2.453 min Scan# 224
Delta R.T. -0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47


Tgt Ion:142 Resp: 90407
Ion Ratio Lower Upper
142 100
127 47.7 41.6 62.4
141 14.2 11.6 17.4








#11
Tert butyl alcohol
Concen: 229.992 ug/l
RT: 2.959 min Scan# 307
Delta R.T. -0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 59 Resp: 107775
Ion Ratio Lower Upper
59 100
57 10.3 6.0 9.0#




#12
1,1-Dichloroethene
Concen: 51.872 ug/l
RT: 2.319 min Scan# 202
Delta R.T. -0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47



Tgt Ion: 96 Resp: 62396
Ion Ratio Lower Upper
96 100
61 163.9 144.6 216.8
98 67.2 50.4 75.6









#13
Acrolein
Concen: 331.201 ug/l
RT: 2.233 min Scan# 188
Delta R.T. 0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 56 Resp: 99566
Ion Ratio Lower Upper
56 100
55 70.4 57.1 85.7



#14
Allyl chloride
Concen: 48.954 ug/l
RT: 2.660 min Scan# 258
Delta R.T. 0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47

Tgt Ion: 41 Resp: 105326
Ion Ratio Lower Upper
41 100
39 73.2 55.8 83.6
76 35.9 24.3 36.5









#15
Acrylonitrile
Concen: 252.867 ug/l
RT: 3.062 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 53 Resp: 243343
Ion Ratio Lower Upper
53 100
52 83.5 66.0 99.0
51 36.7 29.8 44.6



#16
Acetone
Concen: 245.702 ug/l
RT: 2.380 min Scan# 212
Delta R.T. -0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47

Tgt Ion: 43 Resp: 216845
Ion Ratio Lower Upper
43 100
58 31.3 21.9 32.9





#17

Carbon Disulfide

Concen: 45.171 ug/l

RT: 2.514 min Scan# 21

Delta R.T. 0.006 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 76 Resp: 156601

Ion Ratio Lower Upper

76 100

78 9.6 7.0 10.6



#18

Methyl Acetate

Concen: 51.969 ug/l

RT: 2.703 min Scan# 265

Delta R.T. -0.000 min

Lab File: VX038213.D




Acq: 10 Oct 2023 19:47


Tgt Ion: 43 Resp: 196685

Ion Ratio Lower Upper

43 100

74 25.8 17.8 26.8









#19
Methyl tert-butyl Ether
Concen: 50.743 ug/l
RT: 3.111 min Scan# 311
Delta R.T. 0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 73 Resp: 243093
Ion Ratio Lower Upper
73 100
57 21.4 17.3 25.9



#20
Methylene Chloride
Concen: 46.822 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47


Tgt Ion: 84 Resp: 77727
Ion Ratio Lower Upper
84 100
49 117.8 103.6 155.4
51 35.5 33.8 50.6
86 65.3 49.4 74.0








#21
trans-1,2-Dichloroethene
Concen: 50.070 ug/l
RT: 3.093 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 96 Resp: 74571
Ion Ratio Lower Upper
96 100
61 131.9 119.2 178.8
98 64.8 48.7 73.1




#22
Diisopropyl ether
Concen: 49.881 ug/l
RT: 3.757 min Scan# 438
Delta R.T. -0.007 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47



Tgt Ion: 45 Resp: 224928
Ion Ratio Lower Upper
45 100
43 85.2 69.7 104.5
87 30.5 20.8 31.2
59 13.5 9.0 13.4#









#23
Vinyl Acetate
Concen: 252.987 ug/l
RT: 3.721 min Scan# 41
Delta R.T. -0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 836590
Ion Ratio Lower Upper
43 100
86 11.4 7.8 11.8



#24
1,1-Dichloroethane
Concen: 49.896 ug/l
RT: 3.611 min Scan# 414
Delta R.T. 0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47

Tgt Ion: 63 Resp: 136544
Ion Ratio Lower Upper
63 100
98 6.6 3.3 9.8
100 4.5 2.1 6.5









#25
2-Butanone
Concen: 247.587 ug/l
RT: 4.556 min Scan# 569
Delta R.T. 0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 352557
Ion Ratio Lower Upper
43 100
72 29.5 19.4 29.2#



#26
2,2-Dichloropropane
Concen: 47.099 ug/l
RT: 4.477 min Scan# 556
Delta R.T. -0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47

Tgt Ion: 77 Resp: 100631
Ion Ratio Lower Upper
77 100
97 23.4 10.8 32.3









#27
 cis-1,2-Dichloroethene
 Concen: 50.209 ug/l
 RT: 4.489 min Scan# 511
 Delta R.T. 0.000 min
 Lab File: VX038213.D
 Acq: 10 Oct 2023 19:47




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 96 Resp: 89058
 Ion Ratio Lower Upper
 96 100
 61 140.0 0.0 312.2
 98 64.7 0.0 129.0



#28
 Bromochloromethane
 Concen: 52.690 ug/l
 RT: 4.897 min Scan# 625
 Delta R.T. 0.000 min
 Lab File: VX038213.D
 Acq: 10 Oct 2023 19:47

Tgt Ion: 49 Resp: 65092
 Ion Ratio Lower Upper
 49 100
 129 2.0 0.0 3.8
 130 82.5 57.5 86.3





#29

Tetrahydrofuran

Concen: 246.039 ug/l

RT: 5.001 min Scan# 642

Delta R.T. 0.000 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 42 Resp: 228102

Ion Ratio Lower Upper

42 100

72 49.8 34.2 51.4

71 47.1 32.2 48.2



#30

Chloroform

Concen: 51.929 ug/l

RT: 5.093 min Scan# 657

Delta R.T. -0.000 min

Lab File: VX038213.D




Acq: 10 Oct 2023 19:47


Tgt Ion: 83 Resp: 155398

Ion Ratio Lower Upper

83 100


85 63.0 50.3 75.5








#31
Cyclohexane
Concen: 48.194 ug/l
RT: 5.477 min Scan# 719
Delta R.T. 0.006 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 56 Resp: 110417
Ion Ratio Lower Upper
56 100
69 30.7 23.4 35.2
84 89.1 65.9 98.9




#32
1,1,1-Trichloroethane
Concen: 51.753 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47



Tgt Ion: 97 Resp: 132653
Ion Ratio Lower Upper
97 100
99 63.7 51.1 76.7
61 44.3 36.4 54.6





#33

1,2-Dichloroethane-d4

Concen: 47.089 ug/l

RT: 5.952 min Scan# 799

Delta R.T. -0.006 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 65 Resp: 99578

Ion Ratio Lower Upper

65 100

67 52.2 0.0 100.2



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. 0.000 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47




Tgt Ion: 114 Resp: 220502


Ion Ratio Lower Upper

114 100

63 19.6 0.0 45.0

88 15.6 0.0 36.6









#35
Dibromofluoromethane
Concen: 49.685 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:113 Resp: 78943
Ion Ratio Lower Upper
113 100
111 103.4 82.9 124.3
192 18.4 13.3 19.9



#36
1,1-Dichloropropene
Concen: 49.356 ug/l
RT: 5.690 min Scan# 755
Delta R.T. 0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47

Tgt Ion: 75 Resp: 108140
Ion Ratio Lower Upper
75 100
110 35.6 17.0 50.9
77 30.2 24.7 37.1





#37

Ethyl Acetate

Concen: 50.810 ug/l

RT: 4.715 min Scan# 595

Delta R.T. -0.000 min


Lab File: VX038213.D

Acq: 10 Oct 2023 19:47

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion: 43 Resp: 132213

Ion Ratio Lower Upper

43 100

61 14.6 10.6 16.0

70 11.1 8.3 12.5



#38

Carbon Tetrachloride

Concen: 52.510 ug/l

RT: 5.678 min Scan# 753

Delta R.T. -0.000 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47




Tgt Ion: 117 Resp: 113179


Ion Ratio Lower Upper

117 100

119 91.9 78.2 117.2

121 30.1 24.4 36.6





#39

Methylcyclohexane

Concen: 49.327 ug/l

RT: 7.379 min Scan# 1032

Delta R.T. -0.006 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 126590

Ion Ratio Lower Upper

83 100

55 72.2 64.6 97.0

98 45.1 36.7 55.1



#40

Benzene

Concen: 51.614 ug/l

RT: 6.038 min Scan# 812

Delta R.T. -0.000 min

Lab File: VX038213.D




Acq: 10 Oct 2023 19:47


Tgt Ion: 78 Resp: 316301

Ion Ratio Lower Upper

78 100


77 24.1 19.3 28.9








#41
Methacrylonitrile
Concen: 53.023 ug/l
RT: 4.916 min Scan# 61
Delta R.T. -0.006 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 41	Resp: 68636
Ion Ratio	Lower Upper
41 100	
39 57.1	45.2 67.8
67 78.8	52.6 78.8
52 33.5	22.7 34.1



#42
1,2-Dichloroethane
Concen: 50.876 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47


Tgt Ion: 62	Resp: 124565
Ion Ratio	Lower Upper
62 100	
98 8.3	0.0 16.0








#43
Isopropyl Acetate
Concen: 51.360 ug/l
RT: 6.342 min Scan# 862
Delta R.T. 0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 43 Resp: 199511
Ion Ratio Lower Upper
43 100
61 20.0 15.2 22.8
87 14.4 9.8 14.6



#44
Trichloroethene
Concen: 52.158 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47

Tgt Ion:130 Resp: 83599
Ion Ratio Lower Upper
130 100
95 98.9 0.0 204.4









#45
1,2-Dichloropropane
Concen: 51.781 ug/l
RT: 7.428 min Scan# 1040
Delta R.T. -0.006 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 63 Resp: 82129
Ion Ratio Lower Upper
63 100
65 31.4 24.3 36.5



#46
Dibromomethane
Concen: 51.352 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. -0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47

Tgt Ion: 93 Resp: 63170
Ion Ratio Lower Upper
93 100
95 83.2 66.6 100.0
174 92.1 69.4 104.0









#47
Bromodichloromethane
Concen: 54.128 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 118451
Ion Ratio Lower Upper
83 100
85 63.9 49.8 74.6
127 7.7 7.1 10.7



#48
Methyl methacrylate
Concen: 52.268 ug/l
RT: 7.690 min Scan# 1083
Delta R.T. -0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47

Tgt Ion: 41 Resp: 96927
Ion Ratio Lower Upper
41 100
69 90.3 59.2 88.8#
39 66.5 47.7 71.5





#49

1,4-Dioxane

Concen: 1003.506 ug/l

RT: 7.659 min Scan# 1

Delta R.T. 0.000 min


Lab File: VX038213.D

Acq: 10 Oct 2023 19:47




Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion:	88	Resp:	45857
Ion	Ratio	Lower	Upper
88	100		
43	32.6	0.0	0.0#
58	65.1	60.1	90.1



#50

Toluene-d8

Concen: 50.280 ug/l




RT: 8.647 min Scan# 1240


Delta R.T. -0.006 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47

Tgt Ion:	98	Resp:	305269
Ion	Ratio	Lower	Upper
98	100		
100	63.4	53.1	79.7









#51
4-Methyl-2-Pentanone
Concen: 267.535 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 716830
Ion Ratio Lower Upper
43 100
58 40.1 30.4 45.6



#52
Toluene
Concen: 53.388 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47

Tgt Ion: 92 Resp: 213118
Ion Ratio Lower Upper
92 100
91 176.6 134.2 201.2





#53

t-1,3-Dichloropropene

Concen: 54.769 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 75 Resp: 128653

Ion Ratio Lower Upper

75 100

77 31.9 25.7 38.5



#54

cis-1,3-Dichloropropene

Concen: 53.469 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47




Tgt Ion: 75 Resp: 131959


Ion Ratio Lower Upper

75 100

77 31.2 25.3 37.9

39 49.7 46.7 70.1





#55

1,1,2-Trichloroethane

Concen: 53.659 ug/l

RT: 9.153 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion: 97 Resp: 91620





Ion Ratio Lower Upper

97 100

83 85.4 65.8 98.8

85 56.7 41.6 62.4

99 64.0 47.8 71.8



#56

Ethyl methacrylate

Concen: 54.301 ug/l

RT: 9.116 min Scan# 1317

Delta R.T. 0.000 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47




Tgt Ion: 69 Resp: 142154


Ion Ratio Lower Upper

69 100

41 61.2 56.8 85.2

39 40.1 32.0 48.0





#57

1,3-Dichloropropane

Concen: 53.624 ug/l

RT: 9.311 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 76 Resp: 148087

Ion Ratio Lower Upper

76 100

78 32.4 26.1 39.1



#58

2-Chloroethyl Vinyl ether

Concen: 250.244 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.000 min

Lab File: VX038213.D




Acq: 10 Oct 2023 19:47


Tgt Ion: 63 Resp: 321569

Ion Ratio Lower Upper

63 100

106 28.5 22.5 33.7









#59
2-Hexanone
Concen: 267.401 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. -0.006 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 575990
Ion Ratio Lower Upper
43 100
58 55.8 26.2 78.5



#60
Dibromochloromethane
Concen: 57.170 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. -0.006 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47

Tgt Ion: 129 Resp: 95536
Ion Ratio Lower Upper
129 100
127 77.8 38.8 116.4









#61
1,2-Dibromoethane
Concen: 54.492 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:107 Resp: 100617
Ion Ratio Lower Upper
107 100
109 94.1 75.2 112.8



#62
4-Bromofluorobenzene
Concen: 52.323 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47


Tgt Ion: 95 Resp: 120434
Ion Ratio Lower Upper
95 100
174 76.5 0.0 135.6
176 75.8 0.0 131.0








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:117 Resp: 222221
Ion Ratio Lower Upper
117 100
82 56.7 46.2 69.4
119 33.0 25.5 38.3



#64
Tetrachloroethene
Concen: 51.104 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47

Tgt Ion:164 Resp: 73518
Ion Ratio Lower Upper
164 100
166 128.0 102.7 154.1
129 95.2 80.4 120.6
131 95.2 76.5 114.7









#65
Chlorobenzene
Concen: 50.565 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:112 Resp: 235130
Ion Ratio Lower Upper
112 100
114 32.4 25.7 38.5



#66
1,1,1,2-Tetrachloroethane
Concen: 55.658 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47

Tgt Ion:131 Resp: 87518
Ion Ratio Lower Upper
131 100
133 95.1 47.1 141.4
119 64.7 32.2 96.6





#67

Ethyl Benzene

Concen: 51.541 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.000 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 91 Resp: 427806

Ion Ratio Lower Upper

91 100

106 32.0 24.9 37.3



#68

m/p-Xylenes

Concen: 103.879 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.006 min

Lab File: VX038213.D




Acq: 10 Oct 2023 19:47


Tgt Ion:106 Resp: 333394

Ion Ratio Lower Upper

106 100

91 204.5 165.5 248.3





#69

o-Xylene

Concen: 51.487 ug/l

RT: 10.640 min Scan# 1567

Delta R.T. 0.000 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:106 Resp: 161277

Ion Ratio Lower Upper

106 100

91 217.7 108.9 326.8



#70

Styrene

Concen: 53.299 ug/l

RT: 10.653 min Scan# 1569

Delta R.T. -0.006 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47




Tgt Ion:104 Resp: 271890


Ion Ratio Lower Upper

104 100

78 51.6 43.6 65.4

103 56.0 43.6 65.4





#71

Bromoform

Concen: 48.523 ug/l

RT: 10.799 min Scan# 1593

Delta R.T. -0.000 min


Lab File: VX038213.D

Acq: 10 Oct 2023 19:47




Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion:	173	Resp:	70650
Ion Ratio	Lower	Upper	
173	100		
175	49.3	24.1	72.4
254	0.1	0.0	0.0#



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l




RT: 12.024 min Scan# 1794


Delta R.T. 0.000 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47

Tgt Ion:	152	Resp:	129908
Ion Ratio	Lower	Upper	
152	100		
115	85.7	43.3	129.9
150	171.2	0.0	345.0





#73

Isopropylbenzene

Concen: 49.436 ug/l

RT: 10.963 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:105 Resp: 430690

Ion Ratio Lower Upper

105 100

120 26.1 12.9 38.6



#74

N-ethyl acetate

Concen: 49.835 ug/l

RT: 10.842 min Scan# 1600

Delta R.T. -0.000 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47

Tgt Ion: 43 Resp: 185403




Ion Ratio Lower Upper


43 100

70 43.4 32.9 49.3

55 25.0 19.3 28.9

61 24.7 18.8 28.2





#75

1,1,2,2-Tetrachloroethane

Concen: 50.234 ug/l

RT: 11.213 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 160587

Ion Ratio Lower Upper

83 100

131 9.7 5.3 15.8

85 64.5 32.0 96.2



#76

1,2,3-Trichloropropane

Concen: 58.497 ug/l

RT: 11.238 min Scan# 1665

Delta R.T. -0.000 min

Lab File: VX038213.D




Acq: 10 Oct 2023 19:47


Tgt Ion: 75 Resp: 172760

Ion Ratio Lower Upper

75 100

77 32.8 19.1 57.5





#77

Bromobenzene

Concen: 49.654 ug/l

RT: 11.195 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:156 Resp: 105612

Ion Ratio Lower Upper

156 100

77 159.0 88.4 265.3

158 97.1 48.0 144.0



#78

n-propylbenzene

Concen: 48.982 ug/l

RT: 11.305 min Scan# 1676

Delta R.T. -0.000 min

Lab File: VX038213.D




Acq: 10 Oct 2023 19:47


Tgt Ion: 91 Resp: 521787

Ion Ratio Lower Upper

91 100

120 23.0 11.5 34.4









#79
2-Chlorotoluene
Concen: 47.923 ug/l
RT: 11.366 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 302754
Ion Ratio Lower Upper
91 100
126 34.2 16.3 48.8



#80
1,3,5-Trimethylbenzene
Concen: 50.024 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47

Tgt Ion: 105 Resp: 375987
Ion Ratio Lower Upper
105 100
120 46.8 24.1 72.3









#81
trans-1,4-Dichloro-2-butene
Concen: 44.101 ug/l
RT: 11.018 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 44081
Ion Ratio Lower Upper
75 100
53 104.3 90.5 135.7
89 46.6 36.9 55.3



#82
4-Chlorotoluene
Concen: 48.502 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.006 min
Lab File: VX038213.D
Acq: 10 Oct 2023 19:47

Tgt Ion: 91 Resp: 362584
Ion Ratio Lower Upper
91 100
126 30.2 14.2 42.6





#83

tert-Butylbenzene

Concen: 51.160 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:119 Resp: 375735

Ion Ratio Lower Upper

119 100

91 59.6 31.2 93.6

134 21.7 11.7 35.0



#84

1,2,4-Trimethylbenzene

Concen: 50.198 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX038213.D




Acq: 10 Oct 2023 19:47


Tgt Ion:105 Resp: 376188

Ion Ratio Lower Upper

105 100

120 43.7 21.2 63.6









#85
 sec-Butylbenzene
 Concen: 50.512 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX038213.D
 Acq: 10 Oct 2023 19:47




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:105 Resp: 480551
 Ion Ratio Lower Upper
 105 100
 134 19.6 9.8 29.4



#86
 p-Isopropyltoluene
 Concen: 50.973 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX038213.D
 Acq: 10 Oct 2023 19:47

Tgt Ion:119 Resp: 400103
 Ion Ratio Lower Upper
 119 100
 134 25.9 12.8 38.3
 91 25.3 13.1 39.3





#87

1,3-Dichlorobenzene

Concen: 48.594 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min


Lab File: VX038213.D

Acq: 10 Oct 2023 19:47

Instrument :

MSVOA_X

ClientSampleId :






Tgt Ion:146 Resp: 207600

Ion Ratio Lower Upper

146 100

111 40.8 22.1 66.1

148 63.8 31.8 95.3



#88

1,4-Dichlorobenzene

Concen: 48.518 ug/l

RT: 12.043 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47




Tgt Ion:146 Resp: 211612


Ion Ratio Lower Upper

146 100

111 40.4 21.3 63.7

148 63.3 31.8 95.4





#89

n-Butylbenzene

Concen: 49.290 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 91 Resp: 385306

Ion Ratio Lower Upper

91 100

92 51.5 27.4 82.0

134 24.8 12.6 37.8



#90

Hexachloroethane

Concen: 52.041 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX038213.D




Acq: 10 Oct 2023 19:47


Tgt Ion: 117 Resp: 67727

Ion Ratio Lower Upper

117 100

201 71.5 32.9 98.6









#91
 1,2-Dichlorobenzene
 Concen: 49.890 ug/l
 RT: 12.335 min Scan# 1845
 Delta R.T. 0.000 min
 Lab File: VX038213.D
 Acq: 10 Oct 2023 19:47




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:146 Resp: 204872
 Ion Ratio Lower Upper
 146 100
 111 41.7 22.4 67.0
 148 63.4 31.6 95.0



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 50.077 ug/l
 RT: 12.939 min Scan# 1944
 Delta R.T. -0.006 min
 Lab File: VX038213.D
 Acq: 10 Oct 2023 19:47

Tgt Ion: 75 Resp: 38706
 Ion Ratio Lower Upper
 75 100
 155 81.3 36.1 108.5
 157 104.7 44.9 134.7





#93

1,2,4-Trichlorobenzene

Concen: 48.188 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:180 Resp: 132722

Ion Ratio Lower Upper

180 100

182 95.2 48.3 144.8

145 32.1 18.0 54.0



#94

Hexachlorobutadiene

Concen: 48.429 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.000 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47




Tgt Ion:225 Resp: 53686


Ion Ratio Lower Upper

225 100

223 63.6 31.1 93.2

227 65.3 32.5 97.5





#95

Naphthalene

Concen: 48.235 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:128 Resp: 462655

Ion Ratio Lower Upper

128 100

127 12.9 10.3 15.5

129 10.9 8.8 13.2



#96

1,2,3-Trichlorobenzene

Concen: 48.551 ug/l

RT: 13.963 min Scan# 2112

Delta R.T. -0.000 min

Lab File: VX038213.D

Acq: 10 Oct 2023 19:47

Tgt Ion:180 Resp: 129480

Ion Ratio Lower Upper

180 100

182 95.3 46.9 140.6

145 32.8 17.8 53.5

