

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101023\
 Data File : VX038191.D
 Acq On : 10 Oct 2023 11:18
 Operator : JC/MD
 Sample : 04723-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-7

Quant Time: Oct 11 01:35:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	115393	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	205272	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	199653	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100986	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	87228	46.997	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.000%
35) Dibromofluoromethane	5.385	113	68108	46.047	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.100%
50) Toluene-d8	8.647	98	256529	45.387	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.780%#
62) 4-Bromofluorobenzene	11.079	95	94258	43.989	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.980%
Target Compounds						Qvalue
30) Chloroform	5.099	83	3867	1.472	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101023\
Data File : VX038191.D
Acq On : 10 Oct 2023 11:18
Operator : JC/MD
Sample : 04723-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-7

Quant Time: Oct 11 01:35:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

