

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX101023\  
 Data File : VX038192.D  
 Acq On : 10 Oct 2023 11:41  
 Operator : JC/MD  
 Sample : 04782-01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 8 Sample Multiplier: 1

**Instrument :**  
 MSVOA\_X  
**ClientSampleId :**  
 GW-2-10-13

Quant Time: Oct 11 01:35:31 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X100523W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Oct 05 15:27:13 2023  
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
-----						
Internal Standards						
1) Pentafluorobenzene	5.556	168	115271	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	203875	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	196439	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100357	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	85198	45.952	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.900%
35) Dibromofluoromethane	5.385	113	66983	45.596	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.200%
50) Toluene-d8	8.647	98	250427	44.611	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.220%#
62) 4-Bromofluorobenzene	11.079	95	91258	42.881	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.760%

Target Compounds Qvalue  
 -----

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX101023\  
 Data File : VX038192.D  
 Acq On : 10 Oct 2023 11:41  
 Operator : JC/MD  
 Sample : 04782-01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 8 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 GW-2-10-13

Quant Time: Oct 11 01:35:31 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X100523W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu Oct 05 15:27:13 2023  
 Response via : Initial Calibration

