

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101023\
 Data File : VX038201.D
 Acq On : 10 Oct 2023 15:09
 Operator : JC/MD
 Sample : 04782-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-4-10-13

Quant Time: Oct 11 01:38:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	106739	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	190710	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	190904	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94183	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79284	46.180	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.360%
35) Dibromofluoromethane	5.391	113	62429	45.430	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.860%
50) Toluene-d8	8.653	98	243223	46.318	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.640%
62) 4-Bromofluorobenzene	11.079	95	88690	44.551	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.100%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.129	130	855	0.617	ug/l	79
52) Toluene	8.720	92	147296	42.663	ug/l	92
64) Tetrachloroethene	9.275	164	2731	2.210	ug/l	94
67) Ethyl Benzene	10.195	91	4964	0.696	ug/l	97
68) m/p-Xylenes	10.299	106	6007	2.179	ug/l	99
69) o-Xylene	10.646	106	1217	0.452	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101023\
Data File : VX038201.D
Acq On : 10 Oct 2023 15:09
Operator : JC/MD
Sample : 04782-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-4-10-13

Quant Time: Oct 11 01:38:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

