

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101023\
 Data File : VX038206.D
 Acq On : 10 Oct 2023 17:05
 Operator : JC/MD
 Sample : 04782-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-5-23-26

Quant Time: Oct 11 01:40:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	104416	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	189723	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	187042	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92626	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78341	46.646	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.300%
35) Dibromofluoromethane	5.391	113	62447	45.679	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.360%
50) Toluene-d8	8.647	98	235683	45.116	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.240%#
62) 4-Bromofluorobenzene	11.079	95	85707	43.277	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.560%
Target Compounds						
					Qvalue	
30) Chloroform	5.093	83	9034	3.801	ug/l	90
64) Tetrachloroethene	9.275	164	1139	0.941	ug/l #	86
68) m/p-Xylenes	10.305	106	1456	0.539	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101023\
Data File : VX038206.D
Acq On : 10 Oct 2023 17:05
Operator : JC/MD
Sample : 04782-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-5-23-26

Quant Time: Oct 11 01:40:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

