

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101023\
 Data File : VX038208.D
 Acq On : 10 Oct 2023 17:52
 Operator : JC/MD
 Sample : 04782-17 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-1

Quant Time: Oct 11 01:41:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	105394	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	186005	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	185816	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96958	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	77883	45.943	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.880%
35) Dibromofluoromethane	5.385	113	61901	46.185	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.380%
50) Toluene-d8	8.647	98	230785	45.062	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.120%#
62) 4-Bromofluorobenzene	11.079	95	86142	44.366	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.740%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101023\
Data File : VX038208.D
Acq On : 10 Oct 2023 17:52
Operator : JC/MD
Sample : 04782-17 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-1

Quant Time: Oct 11 01:41:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

