

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101023\
 Data File : VX038209.D
 Acq On : 10 Oct 2023 18:15
 Operator : JC/MD
 Sample : 04782-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB

Quant Time: Oct 11 01:41:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	105161	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	191470	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	189890	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	98168	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	79563	47.038	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.080%
35) Dibromofluoromethane	5.385	113	62736	45.472	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.940%
50) Toluene-d8	8.647	98	235531	44.676	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.360%#
62) 4-Bromofluorobenzene	11.079	95	88074	44.066	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101023\
Data File : VX038209.D
Acq On : 10 Oct 2023 18:15
Operator : JC/MD
Sample : 04782-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB

Quant Time: Oct 11 01:41:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

