

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101023\
 Data File : VX038208.D
 Acq On : 10 Oct 2023 17:52
 Operator : JC/MD
 Sample : 04782-17 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-1

Quant Time: Oct 11 01:41:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	105394	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	186005	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	185816	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96958	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	77883	45.943	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.880%
35) Dibromofluoromethane	5.385	113	61901	46.185	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.380%
50) Toluene-d8	8.647	98	230785	45.062	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.120%#
62) 4-Bromofluorobenzene	11.079	95	86142	44.366	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.740%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101023\
Data File : VX038208.D
Acq On : 10 Oct 2023 17:52
Operator : JC/MD
Sample : 04782-17 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-1

Quant Time: Oct 11 01:41:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 711
Delta R.T. -0.006 min
Lab File: VX038208.D
Acq: 10 Oct 2023 17:52




Instrument :
MSVOA_X
ClientSampleId :
DUP-1


Tgt Ion:168 Resp: 105394
Ion Ratio Lower Upper
168 100
99 56.1 56.6 85.0#



#33
1,2-Dichloroethane-d4
Concen: 45.943 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.000 min
Lab File: VX038208.D
Acq: 10 Oct 2023 17:52

Tgt Ion: 65 Resp: 77883
Ion Ratio Lower Upper
65 100
67 51.0 0.0 100.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX038208.D
Acq: 10 Oct 2023 17:52




Instrument :
MSVOA_X
ClientSampleId :
DUP-1


Tgt Ion:114 Resp: 186005
Ion Ratio Lower Upper
114 100
63 20.2 0.0 45.0
88 15.9 0.0 36.6



#35
Dibromofluoromethane
Concen: 46.185 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX038208.D
Acq: 10 Oct 2023 17:52

Tgt Ion:113 Resp: 61901
Ion Ratio Lower Upper
113 100
111 102.3 82.9 124.3
192 18.9 13.3 19.9









#50
Toluene-d8
Concen: 45.062 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX038208.D
Acq: 10 Oct 2023 17:52




Instrument :
MSVOA_X
ClientSampleId :
DUP-1


Tgt Ion: 98 Resp: 230785
Ion Ratio Lower Upper
98 100
100 63.1 53.1 79.7



#62
4-Bromofluorobenzene
Concen: 44.366 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX038208.D
Acq: 10 Oct 2023 17:52

Tgt Ion: 95 Resp: 86142
Ion Ratio Lower Upper
95 100
174 82.4 0.0 135.6
176 79.3 0.0 131.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX038208.D
Acq: 10 Oct 2023 17:52

Instrument :
MSVOA_X
ClientSampleId :
DUP-1

Tgt Ion:117 Resp: 185816
Ion Ratio Lower Upper
117 100
82 55.5 46.2 69.4
119 32.8 25.5 38.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX038208.D
Acq: 10 Oct 2023 17:52

Tgt Ion:152 Resp: 96958
Ion Ratio Lower Upper
152 100
115 60.7 43.3 129.9
150 157.4 0.0 345.0

