

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101025\
 Data File : VX048105.D
 Acq On : 10 Oct 2025 12:50
 Operator : JC/MD
 Sample : Q3324-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-184-TB-20251006

Quant Time: Oct 10 23:46:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	114051	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	233030	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	242170	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	121909	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	89474	49.286	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.580%
35) Dibromofluoromethane	5.379	113	73304	46.905	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.800%
50) Toluene-d8	8.635	98	253162	46.815	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.640%
62) 4-Bromofluorobenzene	11.061	95	109956	53.577	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101025\
Data File : VX048105.D
Acq On : 10 Oct 2025 12:50
Operator : JC/MD
Sample : Q3324-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-184-TB-20251006

Quant Time: Oct 10 23:46:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

