

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101025\
 Data File : VX048109.D
 Acq On : 10 Oct 2025 14:12
 Operator : JC/MD
 Sample : Q3324-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-184-GW-840-842

Quant Time: Oct 10 23:48:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	120601	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	248505	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	261947	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	136698	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	102294	53.287	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.580%	
35) Dibromofluoromethane	5.373	113	83685	50.213	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.420%	
50) Toluene-d8	8.635	98	278924	48.367	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 96.740%	
62) 4-Bromofluorobenzene	11.067	95	123457	56.409	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 112.820%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.307	50	417	0.254	ug/l #	82
16) Acetone	2.380	43	4907	5.879	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101025\
Data File : VX048109.D
Acq On : 10 Oct 2025 14:12
Operator : JC/MD
Sample : Q3324-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-184-GW-840-842

Quant Time: Oct 10 23:48:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

