

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101025\
 Data File : VX048112.D
 Acq On : 10 Oct 2025 15:13
 Operator : JC/MD
 Sample : Q3324-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-184-GW-860-862

Quant Time: Oct 10 23:49:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	118979	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	249066	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	261922	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	131287	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	102388	54.064	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.120%	
35) Dibromofluoromethane	5.379	113	82913	49.637	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.280%	
50) Toluene-d8	8.635	98	281261	48.663	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.320%	
62) 4-Bromofluorobenzene	11.067	95	122524	55.857	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 111.720%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	4046	4.913	ug/l	91
17) Carbon Disulfide	2.520	76	1577	0.408	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101025\
Data File : VX048112.D
Acq On : 10 Oct 2025 15:13
Operator : JC/MD
Sample : Q3324-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-184-GW-860-862

Quant Time: Oct 10 23:49:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

