

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101025\
 Data File : VX048115.D
 Acq On : 10 Oct 2025 16:15
 Operator : JC/MD
 Sample : Q3311-01DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-218DL

Quant Time: Oct 10 23:50:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	107553	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	220883	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	232642	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	113160	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	87628	51.185	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.380%
35) Dibromofluoromethane	5.373	113	72547	48.973	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.940%
50) Toluene-d8	8.635	98	247795	48.343	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.680%
62) 4-Bromofluorobenzene	11.061	95	104105	53.515	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.040%
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	4.495	96	3812	2.263	ug/l	88
30) Chloroform	5.093	83	2217	0.808	ug/l	96
44) Trichloroethene	7.117	130	6452	4.059	ug/l	95
64) Tetrachloroethene	9.263	164	51425	33.908	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101025\
Data File : VX048115.D
Acq On : 10 Oct 2025 16:15
Operator : JC/MD
Sample : Q3311-01DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-218DL

Quant Time: Oct 10 23:50:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

