

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101025\
 Data File : VX048119.D
 Acq On : 10 Oct 2025 17:37
 Operator : JC/MD
 Sample : Q3311-11 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-6BA

Quant Time: Oct 10 23:52:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	106376	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	218035	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	223748	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	109569	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	84165	49.707	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.420%
35) Dibromofluoromethane	5.379	113	70465	48.189	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.380%
50) Toluene-d8	8.635	98	241021	47.635	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.280%
62) 4-Bromofluorobenzene	11.067	95	101113	52.656	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.320%
Target Compounds						
27) cis-1,2-Dichloroethene	4.483	96	541	0.325	ug/l	77
64) Tetrachloroethene	9.263	164	66891	45.860	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101025\
Data File : VX048119.D
Acq On : 10 Oct 2025 17:37
Operator : JC/MD
Sample : Q3311-11 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-6BA

Quant Time: Oct 10 23:52:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

