

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101025\
 Data File : VX048120.D
 Acq On : 10 Oct 2025 17:57
 Operator : JC/MD
 Sample : Q3311-12 200X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-214D

Quant Time: Oct 10 23:53:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	113288	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	232134	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	241401	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	119289	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	90053	49.939	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.880%
35) Dibromofluoromethane	5.379	113	74875	48.095	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.180%
50) Toluene-d8	8.635	98	259052	48.089	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.180%
62) 4-Bromofluorobenzene	11.061	95	109063	53.347	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.700%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	884	1.127	ug/l #	87
64) Tetrachloroethene	9.263	164	3117	1.981	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101025\
Data File : VX048120.D
Acq On : 10 Oct 2025 17:57
Operator : JC/MD
Sample : Q3311-12 200X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-214D

Quant Time: Oct 10 23:53:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

