



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101025\
Data File : VX048099.D
Acq On : 10 Oct 2025 08:36
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Title : SW846 8260
Last Update : Wed Sep 17 06:39:58 2025



AutoFind: Scans 1635, 1636, 1637; Background Corrected with Scan 1629

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	20.2	9993	PASS
75	95	30	60	51.5	25483	PASS
95	95	100	100	100.0	49448	PASS
96	95	5	9	6.2	3042	PASS
173	174	0.00	2	1.4	470	PASS
174	95	50	100	67.0	33144	PASS
175	174	5	9	7.6	2524	PASS
176	174	95	101	99.9	33096	PASS
177	176	5	9	6.3	2070	PASS