

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101118\  
 Data File : VX005268.D  
 Acq On : 11 Oct 2018 21:47  
 Operator : JC/MD  
 Sample : J5393-06  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 15 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :

Quant Time: Oct 12 07:12:14 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X092618W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Oct 03 04:00:03 2018  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) Pentafluorobenzene      | 5.66  | 168  | 181587   | 50.00 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 6.86  | 114  | 264390   | 50.00 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 10.12 | 117  | 267355   | 50.00 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 12.08 | 152  | 150659   | 50.00 | ug/l  | 0.00     |

| System Monitoring Compounds | R.T.  | QIon | Response | Conc  | Units   | Dev(Min) |
|-----------------------------|-------|------|----------|-------|---------|----------|
| 33) 1,2-Dichloroethane-d4   | 6.07  | 65   | 143367   | 55.27 | ug/l    | 0.00     |
| Spiked Amount               |       |      | Recovery | =     | 110.54% |          |
| 35) Dibromofluoromethane    | 5.50  | 113  | 125469   | 55.99 | ug/l    | 0.00     |
| Spiked Amount               |       |      | Recovery | =     | 111.98% |          |
| 50) Toluene-d8              | 8.72  | 98   | 388120   | 52.10 | ug/l    | 0.00     |
| Spiked Amount               |       |      | Recovery | =     | 104.20% |          |
| 62) 4-Bromofluorobenzene    | 11.14 | 95   | 139305   | 51.91 | ug/l    | 0.00     |
| Spiked Amount               |       |      | Recovery | =     | 103.82% |          |


| Target Compounds               | R.T.  | QIon | Response | Conc      | Units  | Qvalue |
|--------------------------------|-------|------|----------|-----------|--------|--------|
| 6) Chloroethane                | 1.70  | 64   | 825      | Below Cal | #      | 45     |
| 9) 1,1,2-Trichlorotrifluoroet  | 2.39  | 101  | 725      | 0.365     | ug/l # | 90     |
| 11) Tert butyl alcohol         | 3.03  | 59   | 5215     | 7.190     | ug/l # | 77     |
| 12) 1,1-Dichloroethene         | 2.37  | 96   | 591      | 0.300     | ug/l # | 88     |
| 13) Acrolein                   | 2.32  | 56   | 143      | 0.277     | ug/l # | 10     |
| 16) Acetone                    | 2.45  | 43   | 8599     | 5.678     | ug/l # | 89     |
| 20) Methylene Chloride         | 2.85  | 84   | 286      | Below Cal | #      | 75     |
| 31) Cyclohexane                | 5.67  | 56   | 3511     | 1.064     | ug/l # | 1      |
| 36) 1,1-Dichloropropene        | 5.67  | 75   | 12704    | 4.155     | ug/l # | 45     |
| 38) Carbon Tetrachloride       | 5.66  | 117  | 16202    | 5.042     | ug/l # | 19     |
| 44) Trichloroethene            | 7.21  | 130  | 6005     | 2.552     | ug/l   | 83     |
| 49) 1,4-Dioxane                | 7.76  | 88   | 153      | 2.316     | ug/l # | 71     |
| 51) 4-Methyl-2-Pentanone       | 8.71  | 43   | 1784     | 0.484     | ug/l # | 1      |
| 58) 2-Chloroethyl Vinyl ether  | 8.04  | 63   | 20       | 14.252    | ug/l # | 49     |
| 60) Dibromochloromethane       | 9.34  | 129  | 1936     | 0.840     | ug/l # | 10     |
| 64) Tetrachloroethene          | 9.34  | 164  | 2254     | 0.857     | ug/l # | 81     |
| 74) N-amyl acetate             | 10.91 | 43   | 128      | 1.772     | ug/l # | 53     |
| 76) 1,2,3-Trichloropropane     | 11.14 | 75   | 69916    | 20.705    | ug/l # | 35     |
| 81) trans-1,4-Dichloro-2-buten | 11.14 | 75   | 69916    | 56.212    | ug/l # | 10     |
| 84) 1,2,4-Trimethylbenzene     | 11.81 | 105  | 185      | Below Cal | #      | 51     |
| 95) Naphthalene                | 13.83 | 128  | 230      | 0.738     | ug/l # | 69     |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101118\  
Data File : VX005268.D  
Acq On : 11 Oct 2018 21:47  
Operator : JC/MD  
Sample : J5393-06  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 15 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :

Quant Time: Oct 12 07:12:14 2018  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X092618W.M  
Quant Title : SW846 8260  
QLast Update : Wed Oct 03 04:00:03 2018  
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. 0.00 min

Lab File: VX005268.D

Acq: 11 Oct 2018 21:47

Instrument :

MSVOA\_X


ClientSampleId :

Tot Ion:168 Resp: 181587

Ion Ratio Lower Upper

168 100

99 41.8 39.3 58.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.66

80000

60000



40000

20000

0

Time-->

5.60 5.70 5.80



#6

Chloroethane

Concen: Below Cal

RT: 1.70 min Scan# 90

Delta R.T. -0.01 min

Lab File: VX005268.D


Acq: 11 Oct 2018 21:47

Tgt Ion: 64 Resp: 825

Ion Ratio Lower Upper

64 100

66 0.0 23.8 35.6#



Abundance Ion 63.90 (63.60 to 64.60): VX0

Ion 65.90 (65.60 to 66.60): VX0

1.70

1200

1000

800

600


400


200

0

Time-->

1.66 1.68 1.70 1.72 1.74





#9

1,1,2-Trichlorotrifluoroethane

Concen: 0.365 ug/l

RT: 2.39 min Scan# 203

Delta R.T. 0.01 min

Lab File: VX005268.D

Acq: 11 Oct 2018 21:47

Instrument :

MSVOA\_X

ClientSampleId :


Tot Ion:101 Resp: 725

Ion Ratio Lower Upper

101 100

85 59.9 35.9 53.9#




151 80.8 66.6 100.0



Abundance Ion 100.90 (100.60 to 101.60): V

Ion 84.90 (84.60 to 85.60): VX0

Ion 150.80 (150.50 to 151.50): V



#11

Tert butyl alcohol

Concen: 7.190 ug/l

RT: 3.03 min Scan# 307

Delta R.T. -0.05 min

Lab File: VX005268.D


Acq: 11 Oct 2018 21:47

Tgt Ion: 59 Resp: 5215

Ion Ratio Lower Upper



59 100


57 2.1 8.6 12.8#



Abundance Ion 59.00 (58.70 to 59.70): VX0

Ion 57.00 (56.70 to 57.70): VX0





#12

1,1-Dichloroethene

Concen: 0.300 ug/l

RT: 2.37 min Scan# 200

Delta R.T. 0.01 min

Lab File: VX005268.D

Acq: 11 Oct 2018 21:47

Instrument :

MSVOA\_X

ClientSampled :


Tot Ion: 96 Resp: 591

Ion Ratio Lower Upper




96 100

61 165.6 130.2 195.4

98 94.9 53.4 80.0#



Abundance Ion 95.90 (95.60 to 96.60): VX0  
Ion 60.90 (60.60 to 61.60): VX0  
Ion 97.90 (97.60 to 98.60): VX0



#13

Acrolein

Concen: 0.277 ug/l

RT: 2.32 min Scan# 191

Delta R.T. 0.02 min

Lab File: VX005268.D


Acq: 11 Oct 2018 21:47

Tgt Ion: 56 Resp: 143



Ion Ratio Lower Upper


56 100

55 0.0 63.4 95.2#



Abundance Ion 56.00 (55.70 to 56.70): VX0  
Ion 55.00 (54.70 to 55.70): VX0





#16

Acetone

Concen: 5.678 ug/l

RT: 2.45 min Scan# 212

Delta R.T. 0.00 min

Lab File: VX005268.D

Acq: 11 Oct 2018 21:47

Instrument :

MSVOA\_X





ClientSampled :

Tot Ion: 43 Resp: 8599

Ion Ratio Lower Upper

43 100

58 32.7 21.8 32.6#



#20

Methylene Chloride

Concen: Below Cal

RT: 2.85 min Scan# 279

Delta R.T. 0.00 min

Lab File: VX005268.D

Acq: 11 Oct 2018 21:47

Tgt Ion: 84 Resp: 286




Ion Ratio Lower Upper


84 100

49 102.0 92.3 138.5

51 101.3 31.8 47.6#

86 60.9 48.5 72.7





#31

Cyclohexane

Concen: 1.064 ug/l

RT: 5.67 min Scan# 740

Delta R.T. 0.09 min

Lab File: VX005268.D

Acq: 11 Oct 2018 21:47

Instrument :

MSVOA\_X

ClientSampled :


Tot Ion: 56 Resp: 3511

Ion Ratio Lower Upper

56 100

69 185.6 25.9 38.9#




84 172.2 71.9 107.9#



Abundance Ion 56.00 (55.70 to 56.70): VX0

Ion 69.00 (68.70 to 69.70): VX0

Ion 84.00 (83.70 to 84.70): VX0



#33

1,2-Dichloroethane-d4

Concen: 55.269 ug/l

RT: 6.07 min Scan# 806

Delta R.T. 0.00 min

Lab File: VX005268.D


Acq: 11 Oct 2018 21:47

Tgt Ion: 65 Resp: 143367

Ion Ratio Lower Upper



65 100


67 51.2 0.0 107.4



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0









#34  
 1,4-Difluorobenzene  
 Concen: 50.000 ug/l  
 RT: 6.86 min Scan# 936  
 Delta R.T. 0.00 min  
 Lab File: VX005268.D  
 Acq: 11 Oct 2018 21:47

Instrument :  
 MSVOA\_X  
 ClientSampleId :

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 114     | 100   |       |       |
| 63      | 20.7  | 0.0   | 39.8  |
| 88      | 15.5  | 0.0   | 27.0  |




Abundance Ion 113.90 (113.60 to 114.60): V  
 Ion 63.00 (62.70 to 63.70): VX0  
 Ion 87.90 (87.60 to 88.60): VX0






#35  
 Dibromofluoromethane  
 Concen: 55.994 ug/l  
 RT: 5.50 min Scan# 713  
 Delta R.T. 0.00 min  
 Lab File: VX005268.D  
 Acq: 11 Oct 2018 21:47

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 113     | 100   |       |       |
| 111     | 100.5 | 77.0  | 115.4 |
| 192     | 24.6  | 17.0  | 25.6  |



Abundance Ion 112.80 (112.50 to 113.50): V  
 Ion 110.80 (110.50 to 111.50): V  
 Ion 191.70 (191.40 to 192.40): V









#36  
 1,1-Dichloropropene  
 Concen: 4.155 ug/l  
 RT: 5.67 min Scan# 740  
 Delta R.T. -0.13 min  
 Lab File: VX005268.D  
 Acq: 11 Oct 2018 21:47




Instrument :  
 MSVOA\_X  
 ClientSampled :


| Tot Ion   | 75  | Resp  | 12704 |
|-----------|-----|-------|-------|
| Ion Ratio | 100 | Lower | Upper |
| 75        | 100 |       |       |
| 110       | 8.4 | 20.0  | 59.9# |
| 77        | 0.1 | 27.1  | 40.7# |



#38  
 Carbon Tetrachloride  
 Concen: 5.042 ug/l  
 RT: 5.66 min Scan# 739  
 Delta R.T. -0.12 min  
 Lab File: VX005268.D  
 Acq: 11 Oct 2018 21:47

| Tgt Ion   | 117 | Resp  | 16202  |
|-----------|-----|-------|--------|
| Ion Ratio | 100 | Lower | Upper  |
| 117       | 100 |       |        |
| 119       | 7.6 | 75.2  | 112.8# |
| 121       | 0.0 | 22.5  | 33.7#  |





#44

Trichloroethene

Concen: 2.552 ug/l

RT: 7.21 min Scan# 994

Delta R.T. 0.00 min

Lab File: VX005268.D

Acq: 11 Oct 2018 21:47

Instrument :

MSVOA\_X


ClientSampleId :

Tot Ion:130 Resp: 6005

Ion Ratio Lower Upper




130 100

95 109.5 0.0 186.8



Abundance Ion 129.80 (129.50 to 130.50): V

Ion 94.90 (94.60 to 95.60): VX0



#49

1,4-Dioxane

Concen: 2.316 ug/l

RT: 7.76 min Scan# 1084

Delta R.T. 0.01 min

Lab File: VX005268.D

Acq: 11 Oct 2018 21:47


Tgt Ion: 88 Resp: 153

Ion Ratio Lower Upper

88 100

43 48.4 25.1 37.7#



58 47.1 56.2 84.2#




Abundance Ion 88.00 (87.70 to 88.70): VX0

Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0









#50  
Toluene-d8  
Concen: 52.101 ug/l  
RT: 8.72 min Scan# 1241  
Delta R.T. 0.00 min  
Lab File: VX005268.D  
Acq: 11 Oct 2018 21:47




Instrument :  
MSVOA\_X  
ClientSampleId :


Tot Ion: 98 Resp: 388120  
Ion Ratio Lower Upper  
98 100  
100 63.0 51.9 77.9



#51  
4-Methyl-2-Pentanone  
Concen: 0.484 ug/l  
RT: 8.71 min Scan# 1240  
Delta R.T. 0.06 min  
Lab File: VX005268.D  
Acq: 11 Oct 2018 21:47

Tgt Ion: 43 Resp: 1784  
Ion Ratio Lower Upper  
43 100  
58 168.3 30.5 45.7#









#58  
 2-Chloroethyl Vinyl ether  
 Concen: 14.252 ug/l  
 RT: 8.04 min Scan# 1130  
 Delta R.T. -0.27 min  
 Lab File: VX005268.D  
 Acq: 11 Oct 2018 21:47

Instrument :  
 MSVOA\_X  
 ClientSampleId :

Tot Ion: 63 Resp: 20  
 Ion Ratio Lower Upper  
 63 100  
 106 0.0 21.0 31.6#




Abundance Ion 62.90 (62.60 to 63.60): VX004724.D (-1157) (-)  
 Ion 105.90 (105.60 to 106.60): VX004724.D (-1157) (-)






#60  
 Dibromochloromethane  
 Concen: 0.840 ug/l  
 RT: 9.34 min Scan# 1343  
 Delta R.T. -0.24 min  
 Lab File: VX005268.D  
 Acq: 11 Oct 2018 21:47

Tgt Ion: 129 Resp: 1936  
 Ion Ratio Lower Upper  
 129 100  
 127 0.0 39.1 117.2#



Abundance Ion 128.80 (128.50 to 129.50): VX004724.D (-1375) (-)  
 Ion 126.80 (126.50 to 127.50): VX004724.D (-1375) (-)









#62  
4-Bromofluorobenzene  
Concen: 51.909 ug/l  
RT: 11.14 min Scan# 1638  
Delta R.T. 0.00 min  
Lab File: VX005268.D  
Acq: 11 Oct 2018 21:47

Instrument :  
MSVOA\_X  
ClientSampled :

| Tot Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 95      | 100   |       |       |
| 174     | 95.8  | 0.0   | 185.0 |
| 176     | 91.1  | 0.0   | 180.2 |




Abundance Ion 94.90 (94.60 to 95.60): VX0  
Ion 173.80 (173.50 to 174.50): V  
Ion 175.80 (175.50 to 176.50): V






#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.12 min Scan# 1470  
Delta R.T. 0.00 min  
Lab File: VX005268.D  
Acq: 11 Oct 2018 21:47

| Tgt Ion | Ratio | Lower | Upper |
|---------|-------|-------|-------|
| 117     | 100   |       |       |
| 82      | 54.8  | 42.2  | 63.4  |
| 119     | 32.6  | 27.4  | 41.0  |



Abundance Ion 116.90 (116.60 to 117.60): V  
Ion 82.00 (81.70 to 82.70): VX0  
Ion 118.90 (118.60 to 119.60): V





#64

Tetrachloroethene

Concen: 0.857 ug/l

RT: 9.34 min Scan# 1342

Delta R.T. 0.00 min

Lab File: VX005268.D

Acq: 11 Oct 2018 21:47

Instrument :

MSVOA\_X

ClientSampled :

Tot Ion:164 Resp: 2254


Ion Ratio Lower Upper

164 100

166 149.5 98.4 147.6#

129 80.8 76.1 114.1

131 104.0 71.0 106.4






Abundance Ion 163.80 (163.50 to 164.50): V

Ion 165.80 (165.50 to 166.50): V

Ion 128.80 (128.50 to 129.50): V

Ion 130.80 (130.50 to 131.50): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.08 min Scan# 1792

Delta R.T. 0.00 min

Lab File: VX005268.D

Acq: 11 Oct 2018 21:47


Tgt Ion:152 Resp: 150659

Ion Ratio Lower Upper

152 100

115 55.4 40.2 120.6



150 158.2 0.0 351.0




Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V





#74

N-amvl acetate

Concen: 1.772 ug/l

RT: 10.91 min Scan# 1601

Delta R.T. 0.01 min

Lab File: VX005268.D

Acq: 11 Oct 2018 21:47

Instrument :

MSVOA\_X

ClientSampleId :

Tot Ion: 43 Resp: 128



Ion Ratio Lower Upper

43 100

70 0.0 32.6 48.8#

55 35.9 21.6 32.4#

61 0.0 19.1 28.7#





Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 70.00 (69.70 to 70.70): VX0

Ion 55.00 (54.70 to 55.70): VX0

Ion 61.00 (60.70 to 61.70): VX0



#76

1,2,3-Trichloropropane

Concen: 20.705 ug/l

RT: 11.14 min Scan# 1638

Delta R.T. -0.16 min

Lab File: VX005268.D



Acq: 11 Oct 2018 21:47

Tgt Ion: 75 Resp: 69916

Ion Ratio Lower Upper


75 100


77 1.3 21.4 64.3#



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 77.00 (76.70 to 77.70): VX0





#81

trans-1,4-Dichloro-2-butene

Concen: 56.212 ug/l

RT: 11.14 min Scan# 1638

Delta R.T. 0.06 min

Lab File: VX005268.D

Acq: 11 Oct 2018 21:47

Instrument :

MSVOA\_X

ClientSampleId :


Tot Ion: 75 Resp: 69916

Ion Ratio Lower Upper

75 100

53 0.3 80.2 120.4#




89 0.0 37.1 55.7#



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 53.00 (52.70 to 53.70): VX0

Ion 88.90 (88.60 to 89.60): VX0



#95

Naphthalene

Concen: 0.738 ug/l

RT: 13.83 min Scan# 2080

Delta R.T. 0.00 min

Lab File: VX005268.D

Acq: 11 Oct 2018 21:47


Tgt Ion: 128 Resp: 230

Ion Ratio Lower Upper

128 100

127 0.0 10.4 15.6#

129 0.0 8.7 13.1#



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V

