

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX101118\
 Data File : VX005263.D
 Acq On : 11 Oct 2018 19:33
 Operator : JC/MD
 Sample : J5317-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-122D3-20181004

Quant Time: Oct 12 07:45:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	172557	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.87	114	265786	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	265556	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	150776	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	130625	52.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.98%	
35) Dibromofluoromethane	5.51	113	121203	53.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.62%	
50) Toluene-d8	8.72	98	385637	51.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.00%	
62) 4-Bromofluorobenzene	11.14	95	138399	51.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.60%	
Target Compounds						
25) 2-Butanone	4.71	43	5487	2.59	ug/l	# 81
44) Trichloroethene	7.21	130	9633	4.07	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX101118\
Data File : VX005263.D
Acq On : 11 Oct 2018 19:33
Operator : JC/MD
Sample : J5317-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-122D3-20181004

Quant Time: Oct 12 07:45:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

