

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX101118\
 Data File : VX005268.D
 Acq On : 11 Oct 2018 21:47
 Operator : JC/MD
 Sample : J5393-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-123D2-20181008

Quant Time: Oct 12 07:55:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	181587	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	264390	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	267355	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	150659	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	143367	55.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.54%	
35) Dibromofluoromethane	5.50	113	125469	55.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.98%	
50) Toluene-d8	8.72	98	388120	52.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.20%	
62) 4-Bromofluorobenzene	11.14	95	139305	51.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.82%	
Target Compounds						
16) Acetone	2.45	43	8599	5.68	ug/l #	89
44) Trichloroethene	7.21	130	6005	2.55	ug/l	83
64) Tetrachloroethene	9.34	164	2254	0.86	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX101118\
Data File : VX005268.D
Acq On : 11 Oct 2018 21:47
Operator : JC/MD
Sample : J5393-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-123D2-20181008

Quant Time: Oct 12 07:55:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

