

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX101118\
 Data File : VX005269.D
 Acq On : 11 Oct 2018 22:14
 Operator : JC/MD
 Sample : J5393-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE-123D3-20181008

Quant Time: Oct 12 07:56:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 03 04:00:03 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	186559	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	263745	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	255220	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	138834	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	135524	50.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.70%	
35) Dibromofluoromethane	5.50	113	124375	55.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.28%	
50) Toluene-d8	8.72	98	374219	50.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.72%	
62) 4-Bromofluorobenzene	11.14	95	131363	49.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.14%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	2369	1.52	ug/l	92
44) Trichloroethene	7.21	130	1134	0.48	ug/l	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX101118\
Data File : VX005269.D
Acq On : 11 Oct 2018 22:14
Operator : JC/MD
Sample : J5393-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE-123D3-20181008

Quant Time: Oct 12 07:56:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X092618W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 03 04:00:03 2018
Response via : Initial Calibration

