

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101119\
 Data File : VX013011.D
 Acq On : 11 Oct 2019 13:09
 Operator : JC/SP
 Sample : VX1011WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1011WBL01

Quant Time: Oct 11 23:42:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 11 09:48:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	165493	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	271478	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	230338	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	92190	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	102392	49.50	ug/l	-0.01
Spiked Amount	50.000		Recovery	=	99.00%	
35) Dibromofluoromethane	5.48	113	80516	51.67	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.34%	
50) Toluene-d8	8.71	98	313198	51.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.08%	
62) 4-Bromofluorobenzene	11.13	95	108258	45.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.70%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101119\
Data File : VX013011.D
Acq On : 11 Oct 2019 13:09
Operator : JC/SP
Sample : VX1011WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1011WBL01

Quant Time: Oct 11 23:42:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 11 09:48:22 2019
Response via : Initial Calibration

