

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101119\
 Data File : VX013012.D
 Acq On : 11 Oct 2019 13:37
 Operator : JC/SP
 Sample : VX1011MBL01
 Misc : 5.0µ/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1011MBL01

Quant Time: Oct 11 23:46:46 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 11 09:48:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	183496	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	300157	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	260379	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	106465	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	115793	50.49	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.98%	
35) Dibromofluoromethane	5.48	113	88619	51.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.88%	
50) Toluene-d8	8.71	98	355819	52.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.92%	
62) 4-Bromofluorobenzene	11.13	95	121061	46.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.74%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101119\
Data File : VX013012.D
Acq On : 11 Oct 2019 13:37
Operator : JC/SP
Sample : VX1011MBL01
Misc : 5.0µ/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1011MBL01

Quant Time: Oct 11 23:46:46 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 11 09:48:22 2019
Response via : Initial Calibration

