

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101119\
 Data File : VX013022.D
 Acq On : 11 Oct 2019 17:30
 Operator : JC/SP
 Sample : K5261-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-0552

Quant Time: Oct 14 09:13:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 11 09:48:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	178472	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	291230	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	250316	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	101162	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	111793	50.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.24%	
35) Dibromofluoromethane	5.49	113	85779	51.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.64%	
50) Toluene-d8	8.71	98	339005	52.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.00%	
62) 4-Bromofluorobenzene	11.13	95	119120	47.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.06%	
Target Compounds						
3) Chloromethane	1.32	50	1126	0.569	ug/l #	86
4) Vinyl Chloride	1.40	62	8726	4.270	ug/l	96
21) trans-1,2-Dichloroethene	3.16	96	59467	30.763	ug/l	94
27) cis-1,2-Dichloroethene	4.59	96	11298	5.070	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101119\
Data File : VX013022.D
Acq On : 11 Oct 2019 17:30
Operator : JC/SP
Sample : K5261-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S49-0552

Quant Time: Oct 14 09:13:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 11 09:48:22 2019
Response via : Initial Calibration

