

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101119\
 Data File : VX013025.D
 Acq On : 11 Oct 2019 18:40
 Operator : JC/SP
 Sample : K5312-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-0557

Quant Time: Oct 12 04:34:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 11 09:48:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	171388	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	283682	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	238694	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	93650	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	109196	50.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.96%	
35) Dibromofluoromethane	5.49	113	83612	51.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.70%	
50) Toluene-d8	8.71	98	329326	51.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.72%	
62) 4-Bromofluorobenzene	11.14	95	110838	44.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.84%	
Target Compounds						
16) Acetone	2.44	43	1918	1.699	ug/l #	Qvalue 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101119\
Data File : VX013025.D
Acq On : 11 Oct 2019 18:40
Operator : JC/SP
Sample : K5312-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S49-0557

Quant Time: Oct 12 04:34:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 11 09:48:22 2019
Response via : Initial Calibration

