

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
 Data File : VX024711.D
 Acq On : 11 Oct 2021 13:56
 Operator : JC/MD
 Sample : VX1011MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1011MBL01

Quant Time: Oct 11 17:12:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	140886	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	243382	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	238067	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104269	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	105934	53.351	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.700%	
35) Dibromofluoromethane	5.391	113	82931	51.292	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.580%	
50) Toluene-d8	8.653	98	301860	50.725	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.460%	
62) 4-Bromofluorobenzene	11.085	95	117942	50.408	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.820%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
Data File : VX024711.D
Acq On : 11 Oct 2021 13:56
Operator : JC/MD
Sample : VX1011MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1011MBL01

Quant Time: Oct 11 17:12:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

