

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
 Data File : VX024712.D
 Acq On : 11 Oct 2021 14:20
 Operator : JC/MD
 Sample : VX1011WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1011WBL01

Quant Time: Oct 11 17:12:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	144838	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	253668	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	248044	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108215	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	110264	54.016	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.040%
35) Dibromofluoromethane	5.397	113	85292	50.613	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.220%
50) Toluene-d8	8.653	98	311691	50.254	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.500%
62) 4-Bromofluorobenzene	11.085	95	121849	49.966	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
Data File : VX024712.D
Acq On : 11 Oct 2021 14:20
Operator : JC/MD
Sample : VX1011WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1011WBL01

Quant Time: Oct 11 17:12:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

