

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
 Data File : VX024724.D
 Acq On : 11 Oct 2021 19:59
 Operator : JC/MD
 Sample : M4159-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T9-2-TOP-GRAB

Quant Time: Oct 12 06:54:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	137159	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	235056	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	241686	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	106013	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	102126	52.830	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.660%	
35) Dibromofluoromethane	5.397	113	80447	51.518	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.040%	
50) Toluene-d8	8.653	98	300411	52.270	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.540%	
62) 4-Bromofluorobenzene	11.085	95	119735	52.987	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.980%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	17681	12.024	ug/l	96
20) Methylene Chloride	2.794	84	7153	3.890	ug/l #	92
43) Isopropyl Acetate	6.348	43	17442	4.078	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
Data File : VX024724.D
Acq On : 11 Oct 2021 19:59
Operator : JC/MD
Sample : M4159-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T9-2-TOP-GRAB

Quant Time: Oct 12 06:54:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

