

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
 Data File : VX024735.D
 Acq On : 12 Oct 2021 01:49
 Operator : JC/MD
 Sample : PB139767TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB139767TB

Quant Time: Oct 12 08:08:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	131860	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	227825	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	229269	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96534	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	99940	53.777	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.560%	
35) Dibromofluoromethane	5.397	113	76592	50.606	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.220%	
50) Toluene-d8	8.653	98	291105	52.259	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.520%	
62) 4-Bromofluorobenzene	11.085	95	106892	48.804	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.600%	
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	15753	10.370	ug/l	91
20) Methylene Chloride	2.794	84	5424	2.821	ug/l	90
43) Isopropyl Acetate	6.348	43	17718	4.274	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
Data File : VX024735.D
Acq On : 12 Oct 2021 01:49
Operator : JC/MD
Sample : PB139767TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB139767TB

Quant Time: Oct 12 08:08:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

