

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
 Data File : VX024737.D
 Acq On : 12 Oct 2021 02:36
 Operator : JC/MD
 Sample : M4158-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T3-1-TOP-GRAB

Quant Time: Oct 12 08:08:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	131337	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	230157	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	231387	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97462	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	101172	54.657	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.320%	
35) Dibromofluoromethane	5.391	113	78466	51.319	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.640%	
50) Toluene-d8	8.653	98	290157	51.561	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.120%	
62) 4-Bromofluorobenzene	11.085	95	107565	48.614	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.220%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	18130	13.625	ug/l	# 85
20) Methylene Chloride	2.794	84	7267	4.199	ug/l	94
43) Isopropyl Acetate	6.355	43	15744	3.759	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
Data File : VX024737.D
Acq On : 12 Oct 2021 02:36
Operator : JC/MD
Sample : M4158-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T3-1-TOP-GRAB

Quant Time: Oct 12 08:08:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

