

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
 Data File : VX024738.D
 Acq On : 12 Oct 2021 03:00
 Operator : JC/MD
 Sample : M4158-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T3-2-TOP-GRAB

Quant Time: Oct 12 08:09:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	133768	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	234857	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	244371	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107687	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	101742	53.966	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery	= 107.940%		
35) Dibromofluoromethane	5.397	113	80371	51.513	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery	= 103.020%		
50) Toluene-d8	8.652	98	299031	52.074	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery	= 104.140%		
62) 4-Bromofluorobenzene	11.085	95	112718	49.924	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery	= 99.840%		
Target Compounds						Qvalue
16) Acetone	2.385	43	17181	11.942	ug/l	91
20) Methylene Chloride	2.794	84	6681	3.676	ug/l	92
43) Isopropyl Acetate	6.354	43	15695	3.673	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
Data File : VX024738.D
Acq On : 12 Oct 2021 03:00
Operator : JC/MD
Sample : M4158-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T3-2-TOP-GRAB

Quant Time: Oct 12 08:09:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

