

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
 Data File : VX024739.D
 Acq On : 12 Oct 2021 03:23
 Operator : JC/MD
 Sample : M4158-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T3-3-TOP-GRAB

Quant Time: Oct 12 08:09:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	132432	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	234377	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	242587	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103048	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	103169	55.275	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.560%
35) Dibromofluoromethane	5.397	113	80142	51.471	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.940%
50) Toluene-d8	8.653	98	299486	52.260	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.520%
62) 4-Bromofluorobenzene	11.085	95	112451	49.907	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.820%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	17315	12.346	ug/l #	86
20) Methylene Chloride	2.794	84	6842	3.843	ug/l #	78
43) Isopropyl Acetate	6.348	43	16181	3.794	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
Data File : VX024739.D
Acq On : 12 Oct 2021 03:23
Operator : JC/MD
Sample : M4158-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T3-3-TOP-GRAB

Quant Time: Oct 12 08:09:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

