

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
 Data File : VX024741.D
 Acq On : 12 Oct 2021 04:10
 Operator : JC/MD
 Sample : M4158-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T3-1-BOTTOM-GRAB

Quant Time: Oct 12 08:10:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	132475	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	233815	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	235797	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	99807	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	101291	54.251	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 108.500%			
35) Dibromofluoromethane	5.397	113	80603	51.892	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.780%			
50) Toluene-d8	8.653	98	294390	51.494	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 102.980%			
62) 4-Bromofluorobenzene	11.085	95	110693	49.245	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 98.500%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	16589	11.378	ug/l	91
20) Methylene Chloride	2.794	84	6691	3.731	ug/l #	79
43) Isopropyl Acetate	6.355	43	16288	3.828	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
Data File : VX024741.D
Acq On : 12 Oct 2021 04:10
Operator : JC/MD
Sample : M4158-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T3-1-BOTTOM-GRAB

Quant Time: Oct 12 08:10:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

