

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
 Data File : VX024742.D
 Acq On : 12 Oct 2021 04:34
 Operator : JC/MD
 Sample : M4158-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T3-2-BOTTOM-GRAB

Quant Time: Oct 12 08:10:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 23 18:24:34 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	133531	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	234094	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	239112	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104636	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	102702	54.572	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.140%	
35) Dibromofluoromethane	5.391	113	80635	51.851	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.700%	
50) Toluene-d8	8.653	98	299708	52.362	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.720%	
62) 4-Bromofluorobenzene	11.085	95	112579	50.025	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.040%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	17289	12.124	ug/l	93
20) Methylene Chloride	2.794	84	6514	3.563	ug/l #	83
43) Isopropyl Acetate	6.348	43	15997	3.756	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101121\
Data File : VX024742.D
Acq On : 12 Oct 2021 04:34
Operator : JC/MD
Sample : M4158-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T3-2-BOTTOM-GRAB

Quant Time: Oct 12 08:10:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092321W.M
Quant Title : SW846 8260
QLast Update : Thu Sep 23 18:24:34 2021
Response via : Initial Calibration

